

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JACO05
 Lab Code: ACE SDG NO.: Q3835
 Lab File ID: VX048866.D Date Analyzed: 12/15/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:54
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	163589	5.53	267535	6.73	239220	10.03
UPPER LIMIT	327178	6.025	535070	7.233	478440	10.531
LOWER LIMIT	81794.5	5.025	133768	6.233	119610	9.531
EPA SAMPLE NO.						
OWBR-01-170-121025	202756	5.54	340422	6.75	306263	10.04
VX1215WBL01	261636	5.54	438317	6.75	386691	10.04
VX1215WBS01	158819	5.54	258455	6.75	231462	10.04

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.