

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JACO05
 Lab Code: ACE SDG NO.: Q3835
 Lab File ID: VX048822.D Date Analyzed: 12/11/2025
 Instrument ID: MSVOA_X Time Analyzed: 13:02
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	120178	12.00€				
UPPER LIMIT	240356	12.50€				
LOWER LIMIT	60089	11.50€				
EPA SAMPLE NO.						
MW-11A-13.5-121025	130810	12.01				
MW-11B-37.5-121025	112926	12.01				
TB01-121025	135977	12.01				
VX1211WBL01	149114	12.00				
VX1211WBS01	146374	12.00				
VX1211WBSD01	119141	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.