

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JACO05
 Lab Code: ACE SDG NO.: Q3835
 Lab File ID: VX048866.D Date Analyzed: 12/15/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:54
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	113643	12				
UPPER LIMIT	227286	12.5				
LOWER LIMIT	56821.5	11.5				
EPA SAMPLE NO.						
OWBR-01-170-121025	138818	12.01				
VX1215WBL01	170934	12.01				
VX1215WBS01	116904	12.00				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.