

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3839

Analytical Method:

8270E

Client: Tetra Tech, EMI

DataFile:

BF144455.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB170916BS	Benzaldehyde	1700	640	ug/Kg	38				10	133	
	Phenol	1700	1300	ug/Kg	76				62	112	
	bis(2-Chloroethyl)ether	1700	1300	ug/Kg	76				60	101	
	2-Chlorophenol	1700	1400	ug/Kg	82				65	112	
	2-Methylphenol	1700	1400	ug/Kg	82				74	105	
	2,2-oxybis(1-Chloropropane)	1700	1300	ug/Kg	76				51	100	
	Acetophenone	1700	1500	ug/Kg	88				66	98	
	3+4-Methylphenols	1700	1400	ug/Kg	82				58	111	
	N-Nitroso-di-n-propylamine	1700	1400	ug/Kg	82				63	95	
	Hexachloroethane	1700	1400	ug/Kg	82				72	108	
	Nitrobenzene	1700	1500	ug/Kg	88				75	100	
	Isophorone	1700	1400	ug/Kg	82				59	99	
	2-Nitrophenol	1700	1600	ug/Kg	94				75	113	
	2,4-Dimethylphenol	1700	1400	ug/Kg	82				61	130	
	bis(2-Chloroethoxy)methane	1700	1400	ug/Kg	82				66	97	
	2,4-Dichlorophenol	1700	1500	ug/Kg	88				62	107	
	Naphthalene	1700	1300	ug/Kg	76				62	100	
	4-Chloroaniline	1700	710	ug/Kg	42				16	100	
	Hexachlorobutadiene	1700	1300	ug/Kg	76				68	108	
	Caprolactam	1700	1600	ug/Kg	94				67	110	
	4-Chloro-3-methylphenol	1700	1500	ug/Kg	88				74	106	
	2-Methylnaphthalene	1700	1200	ug/Kg	71				60	104	
	Hexachlorocyclopentadiene	3300	2200	ug/Kg	67				45	165	
	2,4,6-Trichlorophenol	1700	1500	ug/Kg	88				59	102	
	2,4,5-Trichlorophenol	1700	1500	ug/Kg	88				74	110	
	1,1-Biphenyl	1700	1500	ug/Kg	88				71	110	
	2-Chloronaphthalene	1700	1300	ug/Kg	76				70	103	
	2-Nitroaniline	1700	1500	ug/Kg	88				66	101	
	Dimethylphthalate	1700	1400	ug/Kg	82				76	101	
	Acenaphthylene	1700	1400	ug/Kg	82				76	111	
	2,6-Dinitrotoluene	1700	1700	ug/Kg	100				78	116	
	3-Nitroaniline	1700	1100	ug/Kg	65				28	100	
	Acenaphthene	1700	1500	ug/Kg	88				73	105	
	2,4-Dinitrophenol	3300	2800	ug/Kg	85				59	137	
	4-Nitrophenol	3300	2900	ug/Kg	88				60	124	
	Dibenzofuran	1700	1400	ug/Kg	82				63	99	
	2,4-Dinitrotoluene	1700	1600	ug/Kg	94				78	115	
	Diethylphthalate	1700	1400	ug/Kg	82				76	104	
	4-Chlorophenyl-phenylether	1700	1300	ug/Kg	76				73	101	
	Fluorene	1700	1400	ug/Kg	82				75	99	
	4-Nitroaniline	1700	1300	ug/Kg	76				64	103	
	4,6-Dinitro-2-methylphenol	1700	1600	ug/Kg	94				76	113	
	N-Nitrosodiphenylamine	1700	1400	ug/Kg	82				71	99	
	4-Bromophenyl-phenylether	1700	1400	ug/Kg	82				66	102	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: <u>Q3839</u>	Analytical Method: <u>8270E</u>
Client: <u>Tetra Tech, EMI</u>	DataFile: <u>BF144455.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB170916BS	Hexachlorobenzene	1700	1400	ug/Kg	82				64	98	
	Atrazine	1700	1500	ug/Kg	88				71	122	
	Pentachlorophenol	3300	2900	ug/Kg	88				56	129	
	Phenanthrene	1700	1300	ug/Kg	76				75	100	
	Anthracene	1700	1300	ug/Kg	76				75	103	
	Carbazole	1700	1300	ug/Kg	76				76	103	
	Di-n-butylphthalate	1700	1500	ug/Kg	88				76	110	
	Fluoranthene	1700	1300	ug/Kg	76				73	105	
	Pyrene	1700	1400	ug/Kg	82				59	103	
	Butylbenzylphthalate	1700	1500	ug/Kg	88				75	120	
	3,3-Dichlorobenzidine	1700	1300	ug/Kg	76				31	94	
	Benzo(a)anthracene	1700	1400	ug/Kg	82				75	104	
	Chrysene	1700	1300	ug/Kg	76				75	103	
	bis(2-Ethylhexyl)phthalate	1700	1600	ug/Kg	94				77	113	
	Di-n-octyl phthalate	1700	1700	ug/Kg	100				76	110	
	Benzo(b)fluoranthene	1700	1500	ug/Kg	88				62	109	
	Benzo(k)fluoranthene	1700	1500	ug/Kg	88				75	108	
	Benzo(a)pyrene	1700	1600	ug/Kg	94				79	109	
	Indeno(1,2,3-cd)pyrene	1700	1500	ug/Kg	88				63	101	
	Dibenz(a,h)anthracene	1700	1500	ug/Kg	88				61	112	
	Benzo(g,h,i)perylene	1700	1500	ug/Kg	88				70	108	
	1,2,4,5-Tetrachlorobenzene	1700	1500	ug/Kg	88				53	101	
	1,4-Dioxane	1700	1200	ug/Kg	71				50	96	
	2,3,4,6-Tetrachlorophenol	1700	1700	ug/Kg	100				71	118	