

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3839

Client ID : SSTDCCC040

Date Analyzed: 12/12/2025

Lab File ID: BP026307.D

Time Analyzed: 10:44

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	311819	7.648	1120150	10.41	664855	14.28
UPPER LIMIT	623638	8.148	2240300	10.913	1329710	14.777
LOWER LIMIT	155910	7.148	560075	9.913	332428	13.777
EPA SAMPLE NO.						
01 COAM6	271202	7.65	1043710	10.41	669968	14.28
02 COAN1MS	227866	7.65	953963	10.43	662338	14.28
03 COAN1MSD	166485	7.65	722158	10.41	504466	14.28
04 COAN1	292953	7.62	1186340	10.39	788592	14.28
05 COAN5	218390	7.65	934074	10.41	637315	14.27
06 COAN6	209247	7.65	945816	10.41	707950	14.27
07 COAN7	271191	7.65	1130140	10.41	780544	14.28

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.