

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3839

Client ID : SSTDCCC040

Date Analyzed: 12/15/2025

Lab File ID: BP026321.D

Time Analyzed: 12:03

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	300235	7.655	1161640	10.41	752543	14.28
UPPER LIMIT	600470	8.155	2323280	10.907	1505090	14.778
LOWER LIMIT	150118	7.155	580820	9.907	376272	13.778
EPA SAMPLE NO.						
01 C0AM5	226025	7.65	842745	10.41	512526	14.27

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.