

Report of Analysis

Client:	Tetra Tech, EMI		Date Collected:	12/09/25
Project:	R37092		Date Received:	12/11/25
Client Sample ID:	C0AT4		SDG No.:	Q3839
Lab Sample ID:	Q3839-22		Matrix:	Water
Analytical Method:	8270E	Level: LOW	% Solid:	0
Sample Wt/Vol:	960 mL	Final Vol: 1000 uL	Test:	SVOCMS Group1
Prep Method :	3510C	Prep Date: 12/12/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS									
100-52-7	Benzaldehyde	4.10	U	1	4.10	10.4	ug/L	12/15/25 17:34	PB170919
108-95-2	Phenol	0.95	U	1	0.95	5.20	ug/L	12/15/25 17:34	PB170919
111-44-4	bis(2-Chloroethyl)ether	0.84	U	1	0.84	5.20	ug/L	12/15/25 17:34	PB170919
95-57-8	2-Chlorophenol	0.60	U	1	0.60	5.20	ug/L	12/15/25 17:34	PB170919
95-48-7	2-Methylphenol	1.20	U	1	1.20	5.20	ug/L	12/15/25 17:34	PB170919
108-60-1	2,2-oxybis(1-Chloropropane)	1.30	U	1	1.30	5.20	ug/L	12/15/25 17:34	PB170919
98-86-2	Acetophenone	0.77	U	1	0.77	5.20	ug/L	12/15/25 17:34	PB170919
65794-96-9	3+4-Methylphenols	1.10	U	1	1.10	10.4	ug/L	12/15/25 17:34	PB170919
621-64-7	n-Nitroso-di-n-propylamine	1.50	U	1	1.50	2.60	ug/L	12/15/25 17:34	PB170919
67-72-1	Hexachloroethane	0.68	U	1	0.68	5.20	ug/L	12/15/25 17:34	PB170919
98-95-3	Nitrobenzene	0.79	U	1	0.79	5.20	ug/L	12/15/25 17:34	PB170919
78-59-1	Isophorone	0.78	U	1	0.78	5.20	ug/L	12/15/25 17:34	PB170919
88-75-5	2-Nitrophenol	1.80	U	1	1.80	5.20	ug/L	12/15/25 17:34	PB170919
105-67-9	2,4-Dimethylphenol	1.90	U	1	1.90	5.20	ug/L	12/15/25 17:34	PB170919
111-91-1	bis(2-Chloroethoxy)methane	0.71	U	1	0.71	5.20	ug/L	12/15/25 17:34	PB170919
120-83-2	2,4-Dichlorophenol	0.54	U	1	0.54	5.20	ug/L	12/15/25 17:34	PB170919
106-47-8	4-Chloroaniline	0.88	U	1	0.88	5.20	ug/L	12/15/25 17:34	PB170919
87-68-3	Hexachlorobutadiene	0.56	U	1	0.56	5.20	ug/L	12/15/25 17:34	PB170919
105-60-2	Caprolactam	1.20	U	1	1.20	10.4	ug/L	12/15/25 17:34	PB170919
59-50-7	4-Chloro-3-methylphenol	0.61	U	1	0.61	5.20	ug/L	12/15/25 17:34	PB170919
77-47-4	Hexachlorocyclopentadiene	3.80	U	1	3.80	10.4	ug/L	12/15/25 17:34	PB170919
88-06-2	2,4,6-Trichlorophenol	0.53	U	1	0.53	5.20	ug/L	12/15/25 17:34	PB170919
95-95-4	2,4,5-Trichlorophenol	0.65	U	1	0.65	5.20	ug/L	12/15/25 17:34	PB170919
92-52-4	1,1-Biphenyl	0.55	U	1	0.55	5.20	ug/L	12/15/25 17:34	PB170919
91-58-7	2-Chloronaphthalene	0.64	U	1	0.64	5.20	ug/L	12/15/25 17:34	PB170919
88-74-4	2-Nitroaniline	1.30	U	1	1.30	5.20	ug/L	12/15/25 17:34	PB170919
131-11-3	Dimethylphthalate	0.64	U	1	0.64	5.20	ug/L	12/15/25 17:34	PB170919
606-20-2	2,6-Dinitrotoluene	0.96	U	1	0.96	5.20	ug/L	12/15/25 17:34	PB170919
99-09-2	3-Nitroaniline	1.10	UQ	1	1.10	5.20	ug/L	12/15/25 17:34	PB170919
51-28-5	2,4-Dinitrophenol	6.20	U	1	6.20	10.4	ug/L	12/15/25 17:34	PB170919
100-02-7	4-Nitrophenol	2.50	U	1	2.50	10.4	ug/L	12/15/25 17:34	PB170919
132-64-9	Dibenzofuran	0.64	U	1	0.64	5.20	ug/L	12/15/25 17:34	PB170919
121-14-2	2,4-Dinitrotoluene	1.30	U	1	1.30	5.20	ug/L	12/15/25 17:34	PB170919
84-66-2	Diethylphthalate	0.72	U	1	0.72	5.20	ug/L	12/15/25 17:34	PB170919
7005-72-3	4-Chlorophenyl-phenylether	0.71	U	1	0.71	5.20	ug/L	12/15/25 17:34	PB170919
100-01-6	4-Nitroaniline	1.60	U	1	1.60	5.20	ug/L	12/15/25 17:34	PB170919
534-52-1	4,6-Dinitro-2-methylphenol	3.00	U	1	3.00	10.4	ug/L	12/15/25 17:34	PB170919
86-30-6	n-Nitrosodiphenylamine	0.60	U	1	0.60	5.20	ug/L	12/15/25 17:34	PB170919
101-55-3	4-Bromophenyl-phenylether	0.42	U	1	0.42	5.20	ug/L	12/15/25 17:34	PB170919
118-74-1	Hexachlorobenzene	0.54	U	1	0.54	5.20	ug/L	12/15/25 17:34	PB170919
1912-24-9	Atrazine	1.10	U	1	1.10	5.20	ug/L	12/15/25 17:34	PB170919
86-74-8	Carbazole	0.75	U	1	0.75	5.20	ug/L	12/15/25 17:34	PB170919

Report of Analysis

Client:	Tetra Tech, EMI		Date Collected:	12/09/25
Project:	R37092		Date Received:	12/11/25
Client Sample ID:	C0AT4		SDG No.:	Q3839
Lab Sample ID:	Q3839-22		Matrix:	Water
Analytical Method:	8270E	Level: LOW	% Solid:	0
Sample Wt/Vol:	960 mL	Final Vol: 1000 uL	Test:	SVOCMS Group1
Prep Method :	3510C	Prep Date: 12/12/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
84-74-2	Di-n-butylphthalate	1.30	U	1	1.30	5.20	ug/L	12/15/25 17:34	PB170919
85-68-7	Butylbenzylphthalate	2.00	U	1	2.00	5.20	ug/L	12/15/25 17:34	PB170919
91-94-1	3,3-Dichlorobenzidine	0.97	U	1	0.97	10.4	ug/L	12/15/25 17:34	PB170919
117-84-0	Di-n-octyl phthalate	2.40	U	1	2.40	10.4	ug/L	12/15/25 17:34	PB170919
95-94-3	1,2,4,5-Tetrachlorobenzene	0.54	U	1	0.54	5.20	ug/L	12/15/25 17:34	PB170919
58-90-2	2,3,4,6-Tetrachlorophenol	0.75	U	1	0.75	5.20	ug/L	12/15/25 17:34	PB170919
SURROGATES									
367-12-4	2-Fluorophenol	47.1			10 - 134	31%	SPK: 150		
13127-88-3	Phenol-d6	26.7			10 - 135	18%	SPK: 150		
4165-60-0	Nitrobenzene-d5	65.4			39 - 138	65%	SPK: 100		
321-60-8	2-Fluorobiphenyl	71.0			52 - 132	71%	SPK: 100		
118-79-6	2,4,6-Tribromophenol	125			44 - 137	83%	SPK: 150		
1718-51-0	Terphenyl-d14	82.3			42 - 152	82%	SPK: 100		
INTERNAL STANDARDS									
		Area Count							
3855-82-1	1,4-Dichlorobenzene-d4	224000							
1146-65-2	Naphthalene-d8	832000							
15067-26-2	Acenaphthene-d10	510000							
1517-22-2	Phenanthrene-d10	1010000							
1719-03-5	Chrysene-d12	1110000							
1520-96-3	Perylene-d12	1250000							
TENTATIVE IDENTIFIED COMPOUNDS									
000057-10-3	n-Hexadecanoic acid	3.70	J			18.0	ug/L		
000057-11-4	Octadecanoic acid	3.10	J			19.4	ug/L		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products