

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3839

Client ID : SSTDCCC040

Date Analyzed: 12/15/2025

Lab File ID: BF144466.D

Time Analyzed: 11:13

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	200298	6.793	793888	8.08	426838	9.83
UPPER LIMIT	400596	7.293	1587780	8.575	853676	10.328
LOWER LIMIT	100149	6.293	396944	7.575	213419	9.328
EPA SAMPLE NO.						
01 COAT5	213047	6.79	827476	8.08	437540	9.83
02 COAT6	220363	6.79	853077	8.08	457944	9.83

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.