

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3839

Client ID : SSTDCCC040

Date Analyzed: 12/15/2025

Lab File ID: BP026321.D

Time Analyzed: 12:03

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	300235	7.655	1161640	10.41	752543	14.28
UPPER LIMIT	600470	8.155	2323280	10.907	1505090	14.778
LOWER LIMIT	150118	7.155	580820	9.907	376272	13.778
EPA SAMPLE NO.						
01 PB170919BL	245455	7.65	922384	10.41	572986	14.28
02 PB170919BS	260280	7.65	1010500	10.41	602119	14.28
03 C0AM7	234076	7.65	872699	10.41	555224	14.27
04 C0AM7MS	213200	7.65	807190	10.41	492568	14.28
05 C0AM7MSD	229645	7.65	876881	10.41	546856	14.28
06 C0B09	190243	7.65	711203	10.41	429739	14.27
07 C0AT4	224255	7.65	832333	10.41	509922	14.28
08 C0AT7	197354	7.65	734748	10.41	449350	14.28
09 C0AT8	232447	7.65	871887	10.41	547997	14.28

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.