

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3839

Client ID: SSTDCCC040

Date Analyzed: 12/15/2025

Lab File ID: BF144466.D

Time Analyzed: 11:13

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	684594	11.31	396863	13.945	375851	15.374
UPPER LIMIT	1369190	11.81	793726	14.445	751702	15.874
LOWER LIMIT	342297	10.81	198432	13.445	187926	14.874
EPA SAMPLE NO.						
01 COAT5	742777	11.31	585600	13.94	421916	15.37
02 COAT6	772123	11.30	615736	13.94	462130	15.37

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.