

**ALLIANCE** 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

**GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY**

ORDER ID: Q3840

MATRIX: Water

METHOD: 7196A

|  | NA | NO | YES |
|--|----|----|-----|
| 1. Blank Contamination - If yes, list compounds and concentrations in each blank:              |    | ✓  |     |
| 2. Matrix Spike Duplicate Recoveries Met Criteria  |    |    | ✓   |
| If not met, list those compounds and their recoveries which fall outside the acceptable range. |    |    |     |
| The Blank Spike met requirements for all compounds.  |    |    |     |
| 3. Sample Duplicate Analysis Met QC Criteria   |    |    | ✓   |
| If not met, list those compounds and their recoveries which fall outside the acceptable range. |    |    |     |
| 4. Digestion Holding Time Met  |    |    | ✓   |
| If not met, list number of days exceeded for each sample:                                      |    |    |     |

ADDITIONAL COMMENTS: As per method, aqueous sample for Hexavalent Chromium analysis should be filtered within 15 minutes of collection time. However, samples were not filtered as per requirement therefore Lab has filtered the samples in-house.

\_\_\_\_\_  
QA REVIEW

\_\_\_\_\_  
Date