

Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: rubina

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB138208

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP116105
Turbidity Calibration std, 20NTU	WP116101
Turbidity Calibration std, 40NTU	WP116100
Turbidity Calibration std, 1NTU	WP116104
Turbidity Calibration std, 5NTU	WP116103
Turbidity Calibration - CCV std, 10 NTU	WP116102
10 NTU Standard 500 ml	W3116

Intercept: 0.2549

Slope: 1.0694

Regression: 0.9998

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.141	-0.11		12/12/2025	09:30
2	CAL2	1	1	1.240	0.92	-7.9	12/12/2025	09:33
3	CAL3	5	1	5.520	4.92	-1.5	12/12/2025	09:36
4	CAL4	10	1	11.600	10.61	6.1	12/12/2025	09:39
5	CAL5	20	1	21.200	19.59	-2.1	12/12/2025	09:42
6	CAL6	40	1	43.100	40.07	0.2	12/12/2025	09:45

Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: rubina

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB138208

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	11.200	10.235	12/12/2025	09:48
2	ICB		1	0.160	-0.089	12/12/2025	09:51
3	CCV1	10	1	10.200	9.300	12/12/2025	09:54
4	CCB1		1	0.158	-0.091	12/12/2025	09:57
5	LB138208BL		1	0.157	-0.092	12/12/2025	10:00
6	Q3841-01		1	1.220	0.902	12/12/2025	10:03
7	Q3841-02		1	1.320	0.996	12/12/2025	10:06
8	Q3841-03		1	1.840	1.482	12/12/2025	10:09
9	Q3841-04		1	0.817	0.526	12/12/2025	10:12
10	Q3841-05		1	0.954	0.654	12/12/2025	10:15
11	Q3841-06		1	0.725	0.440	12/12/2025	10:18
12	Q3841-07		1	0.714	0.429	12/12/2025	10:21
13	Q3841-07DUP		1	0.717	0.432	12/12/2025	10:24
14	CCV2	10	1	10.600	9.674	12/12/2025	10:27
15	CCB2		1	0.156	-0.092	12/12/2025	10:30