

Initial and Continuing Calibration Blank Summary

Client: CHA Companies, Inc.

SDG No.: Q3841

Project: NYSOMH Emergency Service

RunNo.: LB138205

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Residual Chlorine	mg/L	< 0.0500	0.0500	U	0.023	0.1	12/11/2025
Sample ID: CCB1 Residual Chlorine	mg/L	< 0.0500	0.0500	U	0.023	0.1	12/11/2025
Sample ID: CCB2 Residual Chlorine	mg/L	< 0.0500	0.0500	U	0.023	0.1	12/11/2025
Sample ID: CCB3 Residual Chlorine	mg/L	< 0.0500	0.0500	U	0.023	0.1	12/11/2025

Initial and Continuing Calibration Blank Summary

Client: CHA Companies, Inc.

SDG No.: Q3841

Project: NYSOMH Emergency Service

RunNo.: LB138208

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Turbidity	NTU	< 0.5000	0.5000	U	0.15	1.0	12/12/2025
Sample ID: CCB1 Turbidity	NTU	< 0.5000	0.5000	U	0.15	1	12/12/2025
Sample ID: CCB2 Turbidity	NTU	< 0.5000	0.5000	U	0.15	1	12/12/2025