

Laboratory Control Sample Summary

Client: CHA Companies, Inc.

SDG No.: Q3841

Project: NYSOMH Emergency Service

Run No.: LB138205

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138205BS							
Residual Chlorine	mg/L	0.4	0.42		105	1	90-110	12/11/2025