

Initial and Continuing Calibration Verification

Client: JACOBS Engineering Group, Inc.

SDG No.: Q3849

Project: Former Schlumberger STC PTC Site D3868221

RunNo.: LB138212

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	10.1	10	101	90-110	12/01/2025
Chloride	mg/L	3	3	100	90-110	12/01/2025
Fluoride	mg/L	2	2	100	90-110	12/01/2025
Nitrite	mg/L	3	3	100	90-110	12/01/2025
Nitrate	mg/L	2.5	2.5	100	90-110	12/01/2025
Sulfate	mg/L	14.9	15	99	90-110	12/01/2025
Orthophosphate as P	mg/L	5.4	5	108	90-110	12/01/2025
Sample ID: CCV1						
Bromide	mg/L	10.3	10	103	90-110	12/12/2025
Chloride	mg/L	3.1	3	103	90-110	12/12/2025
Fluoride	mg/L	2	2	100	90-110	12/12/2025
Nitrite	mg/L	3.1	3	103	90-110	12/12/2025
Nitrate	mg/L	2.6	2.5	104	90-110	12/12/2025
Sulfate	mg/L	15.1	15	101	90-110	12/12/2025
Orthophosphate as P	mg/L	5.4	5	108	90-110	12/12/2025
Sample ID: CCV2						
Bromide	mg/L	9.2	10	92	90-110	12/12/2025
Chloride	mg/L	2.8	3	93	90-110	12/12/2025
Fluoride	mg/L	1.8	2	90	90-110	12/12/2025
Nitrite	mg/L	2.8	3	93	90-110	12/12/2025
Nitrate	mg/L	2.3	2.5	92	90-110	12/12/2025
Sulfate	mg/L	13.7	15	91	90-110	12/12/2025
Orthophosphate as P	mg/L	4.9	5	98	90-110	12/12/2025
Sample ID: CCV3						
Bromide	mg/L	9.9	10	99	90-110	12/12/2025
Chloride	mg/L	3	3	100	90-110	12/12/2025
Fluoride	mg/L	2	2	100	90-110	12/12/2025
Nitrite	mg/L	3	3	100	90-110	12/12/2025
Nitrate	mg/L	2.5	2.5	100	90-110	12/12/2025
Sulfate	mg/L	14.6	15	97	90-110	12/12/2025
Orthophosphate as P	mg/L	5.4	5	108	90-110	12/12/2025