

Duplicate Sample Summary

Client: JACOBS Engineering Group, Inc. Project: Former Schlumberger STC PTC Site D3868221 Client ID: MW-18B-57.5-121125-FDDUP	SDG No.: Q3849 Sample ID: Q3849-05 Percent Solids for Spike Sample: 0
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Alkalinity	mg/L	+/-20	82.8		84.6		1	2		12/16/2025

Duplicate Sample Summary

Client: JACOBS Engineering Group, Inc. Project: Former Schlumberger STC PTC Site D3868221 Client ID: MW-19B-71.5-121125DUP	SDG No.: Q3849 Sample ID: Q3849-09 Percent Solids for Spike Sample: 0
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TDS	mg/L	+/-5	274		277		1	1.09		12/12/2025

Duplicate Sample Summary

Client: JACOBS Engineering Group, Inc.	SDG No.: Q3849
Project: Former Schlumberger STC PTC Site D3868221	Sample ID: Q3849-09
Client ID: MW-19B-71.5-121125MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Fluoride	mg/L	+/-15	2.40		2.40		1	0		12/12/2025
Nitrate	mg/L	+/-15	2.60		2.60		1	0		12/12/2025
Chloride	mg/L	+/-15	39.1	OR	39.1	OR	1	0		12/12/2025
Sulfate	mg/L	+/-15	93.1	OR	93.0	OR	1	0		12/12/2025
Bromide	mg/L	+/-15	10.5		10.2		1	3		12/12/2025
Nitrite	mg/L	+/-15	3.10		3.00		1	3		12/12/2025
Orthophosphate as P	mg/L	+/-15	1.70		1.60		1	6		12/12/2025

Duplicate Sample Summary

Client: JACOBS Engineering Group, Inc.	SDG No.: Q3849
Project: Former Schlumberger STC PTC Site D3868221	Sample ID: Q3863-01
Client ID: MW-17B-55-121225MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Fluoride	mg/L	+/-15	2.30		2.30		1	0		12/12/2025
Nitrate	mg/L	+/-15	2.40		2.40		1	0		12/12/2025
Nitrite	mg/L	+/-15	2.90		2.90		1	0		12/12/2025
Chloride	mg/L	+/-15	24.0	OR	23.9	OR	1	0		12/12/2025
Bromide	mg/L	+/-15	9.90		9.80		1	1		12/12/2025
Sulfate	mg/L	+/-15	38.5	OR	38.2	OR	1	1		12/12/2025
Orthophosphate as P	mg/L	+/-15	4.30		3.00		1	36	*	12/12/2025