

Laboratory Control Sample Summary

Client:	JACOBS Engineering Group, Inc.	SDG No.:	Q3849
Project:	Former Schlumberger STC PTC Site D3868221	Run No.:	LB138212

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138212BSW							
Bromide	mg/L	10	10.1		101	1	90-110	12/12/2025
Chloride	mg/L	3	3.00		100	1	90-110	12/12/2025
Fluoride	mg/L	2	2.00		100	1	90-110	12/12/2025
Nitrite	mg/L	3	3.00		100	1	90-110	12/12/2025
Nitrate	mg/L	2.5	2.50		100	1	90-110	12/12/2025
Sulfate	mg/L	15	14.9		99	1	90-110	12/12/2025
Orthophosphate as P	mg/L	5	5.40		108	1	90-110	12/12/2025

Laboratory Control Sample Summary

Client: JACOBS Engineering Group, Inc.

SDG No.: Q3849

Project: Former Schlumberger STC PTC Site D3868221

Run No.: LB138213

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138213BS							
TDS	mg/L	100	100		100	1	90-110	12/12/2025

Laboratory Control Sample Summary

Client: JACOBS Engineering Group, Inc.

SDG No.: Q3849

Project: Former Schlumberger STC PTC Site D3868221

Run No.: LB138239

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138239BSW							
Alkalinity	mg/L	50	48.3		97	1	80-120	12/16/2025