

Initial and Continuing Calibration Verification

Client: ALS Environmental

SDG No.: Q3851

Project: ALS Middletown

RunNo.: LB138237

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.99	1	99	90-110	12/16/2025
Sample ID: CCV1 Ammonia as N	mg/L	1	1	100	90-110	12/16/2025
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	12/16/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	12/16/2025
Sample ID: CCV4 Ammonia as N	mg/L	1	1	100	90-110	12/16/2025
Sample ID: CCV5 Ammonia as N	mg/L	1	1	100	90-110	12/16/2025