

Matrix Spike Summary

Client:	ALS Environmental	SDG No.:	Q3851
Project:	ALS Middletown	Sample ID:	Q3851-01
Client ID:	MRS/NRSMS	Percent Solids for Spike Sample:	1.2

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Ammonia as N	mg/Kg	75-125	17800	OR	13600	OR	4000	1	105		12/16/2025

Matrix Spike Summary

Client:	ALS Environmental	SDG No.:	Q3851
Project:	ALS Middletown	Sample ID:	Q3851-01
Client ID:	MRS/NRSMSD	Percent Solids for Spike Sample:	1.2

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Ammonia as N	mg/Kg	75-125	17900	OR	13600	OR	4000	1	108		12/16/2025