

Duplicate Sample Summary

Client: ALS Environmental Project: ALS Middletown Client ID: MRS/NRSDUP	SDG No.: Q3851 Sample ID: Q3851-01 Percent Solids for Spike Sample: 1.2
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Ammonia as N	mg/Kg	+/-20	13600	OR	13700	OR	1	1		12/16/2025
Ammonia as N	mg/Kg	+/-20	13900	D	13900	D	2	0		12/16/2025

Duplicate Sample Summary

Client:	ALS Environmental	SDG No.:	Q3851
Project:	ALS Middletown	Sample ID:	Q3851-01
Client ID:	MRS/NRSMDS	Percent Solids for Spike Sample:	1.2

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/Kg	+/-20	17800	OR	17900	OR	1	1		12/16/2025