

Laboratory Control Sample Summary

Client: ALS Environmental

SDG No.: Q3851

Project: ALS Middletown

Run No.: LB138237

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170957BS							
Ammonia as N	mg/Kg	50	50.2		100	1	90-110	12/16/2025