

Initial and Continuing Calibration Verification

Client: Pacific Commercial Services Inc.

SDG No.: Q3852

Project: Red Hill Pipeline Demo

RunNo.: LB138215

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	6.99	7	100	90-110	12/15/2025
Sample ID: Corrosivity	CCV1	pH	2.01	2.00	101	90-110	12/15/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	12/15/2025