



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: AECOM Technical Services, Inc.

SDG No.: Q3866

Project: NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258

RunNo.: LB138223

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.502	0.5	100	90-110	12/15/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.498	0.5	100	90-110	12/15/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.501	0.5	100	90-110	12/15/2025