

U576328

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : HM

6/2/2015 11:03

Test CN EPA

Instrument ID: CN

Accepted

6/2/2015 11:03

Factor

1043

Bias (y Intercept)

0.005

Coeff. of det.

0.999991

Errors

CN Working Std: WP40980 - WP40985


0.25N NaOH: WP40819

CN CCU Std: WP40987

CN Buffer: WP37745

CN Pyridine: WP37746

CN Chloramphenicol T: WP41001



	Calibrator	Response	Calc. con.	Conc.	Errors	Percentage
1	0.0PPBCN 30.0	0.005	-0.0871	0.0000		98%
2	5.0PPBCN 55.0	0.009	4.5186	5.0000		99%
3	10PPBCN 510.0	0.014	9.9918	10.0000		101%
4	100PPBCN 5100.0	0.102	101.0799	100.0000		99%
5	250PPBCN 5250.0	0.244	249.4157	250.0000		100%
6	500PPBCN 5500.0	0.484	500.0812	500.0000		HM

HM
6/2/15HM
6/2/15

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : LM

6/2/2015 13:51

Test: CN EPA

Sample Id	Result	Dil. 1 +	Response	Errors
ICVPB83727 ICV02	96.581	0.0	0.097	
ICB ICB02	-0.426	0.0	0.004	
CCVWP40987 CCV04	240.653	0.0	0.235	
CCB CCB04	-0.364	0.0	0.004	
PB83727BL PBW07	-0.408	0.0	0.004	
G2390-12 MA4FX9	-0.315	0.0	0.004	
G2453-14 MA4G01	0.121	0.0	0.005	
CCVWP40987 CCV05	240.225	0.0	0.235	
CCB CCB05	-0.109	0.0	0.004	

N	9
Mean	63.995
SD	104.9355
CV%	163.97

CN- Raw Data

1876328

Aquakem v. 7.2AQ1

Results from time period:

Tue Jun 02 13:36:53 2015

Tue Jun 02 13:37:01 2015

IM

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time
S0.0	A	CN EPA	P	-0.0871	µg/l	6/2/2015 11:01:28
S5.0	A	CN EPA	P	4.5186	µg/l	6/2/2015 11:01:29
S10.0	A	CN EPA	P	9.9918	µg/l	6/2/2015 11:01:30
S100.0	A	CN EPA	P	101.0799	µg/l	6/2/2015 11:01:31
S250.0	A	CN EPA	P	249.4157	µg/l	6/2/2015 11:01:32
S500.0	A	CN EPA	P	500.0812	µg/l	6/2/2015 11:01:33
ICVPB83727 ICV02	S	CN EPA	P	96.5807	µg/l	6/2/2015 13:36:53
ICB ICB02	S	CN EPA	P	-0.4259	µg/l	6/2/2015 13:36:54
CCVWP40987 CCV04	S	CN EPA	P	240.6526	µg/l	6/2/2015 13:36:55
CCB CCB04	S	CN EPA	P	-0.3637	µg/l	6/2/2015 13:36:56
PB83727BL PBW07	S	CN EPA	P	-0.4076	µg/l	6/2/2015 13:36:57
G2390-12 MA4FX9	S	CN EPA	P	-0.3148	µg/l	6/2/2015 13:36:58
G2453-14 MA4G01	S	CN EPA	P	0.1206	µg/l	6/2/2015 13:36:59
CCVWP40987 CCV05	S	CN EPA	P	240.2254	µg/l	6/2/2015 13:37:00
CCB CCB05	S	CN EPA	P	-0.1086	µg/l	6/2/2015 13:37:01