

MERCURY ANALYSIS LOGBOOK

[illegible]

MERCURY ANALYSIS LOGBOOK

[illegible]

LB 74458 CN 1


ISM01.3

LB 74458
INSTRUMENT ID: CV1

Linear

 μ Abs.:

50099



A= 0.0000e+000

B= 2.0061e-004

C= -3.0034e-002

Rho= 0.9999902

Accept=Accepted

Slope
y=mx+cept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.008	-0.008	111	0.000	111					15
0.2	0.200	0.210	0.010	1199	0.0 %	1199					0
2.5	2.500	2.505	0.005	12637	0.0 %	12637					0
5.0	5.000	5.003	0.003	25091	0.0 %	25091					0
7.5	7.500	7.469	-0.031	37380	0.0 %	37380					0
10.0	10.000	10.020	0.020	50099	0.0 %	50099					0

NB
1/13/15

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	111	-		0 ISM01.3	PPB	1/14/2015 14:28	Std
	0.2 S0.2	1199	-		0.2 ISM01.3	PPB	1/14/2015 14:30	Std
	2.5 S2.5	12637	-		2.5 ISM01.3	PPB	1/14/2015 14:35	Std
	5 S5	25091	-		5 ISM01.3	PPB	1/14/2015 14:37	Std
	7.5 S7.5	37380	-		7.5 ISM01.3	PPB	1/14/2015 14:39	Std
	10 S10	50099	-		10 ISM01.3	PPB	1/14/2015 14:41	Std
ICV01	ICV01	19656	3.9131	-	ISM01.3	PPB	1/14/2015 14:46	SMPL
ICB01	ICB01	24	-0.0252	-	ISM01.3	PPB	1/14/2015 14:48	SMPL
CCV07	CCV07	24931	4.9713	-	ISM01.3	PPB	1/14/2015 14:50	SMPL
CCB07	CCB07	-16	-0.0332	-	ISM01.3	PPB	1/14/2015 14:53	SMPL
PB81317BL	PBS02	36	-0.0228	-	ISM01.3	PPB	1/14/2015 14:55	SMPL
G1009-01DLX25	MBBTR8	450	0.0602	-	ISM01.3	PPB	1/14/2015 14:57	SMPL
G1009-02DLX25	MBBTR8D	422	0.0546	-	ISM01.3	PPB	1/14/2015 14:59	SMPL
G1009-03DLX25	MBBTR8S	711	0.1126	-	ISM01.3	PPB	1/14/2015 15:01	SMPL
CCV08	CCV08	24362	4.8572	-	ISM01.3	PPB	1/14/2015 15:09	SMPL
CCB08	CCB08	22	-0.0256	-	ISM01.3	PPB	1/14/2015 15:11	SMPL