

MERCURY ANALYSIS LOGBOOK

Date	Case Number	Batch Number	Start Time	BLK	STD1	STD2	STD3	STD4
2/3/15 (7406)	G11493	P03 S1575	13:53	-14	779	10069	20215	2858
2/5/15 (11006)	G1213	P03 S1598	9:10	22	875	10638	21215	31916
	G12214	↓	↓	↓	↓	↓	↓	↓
2/15/15 (7441)	G1212	P03 S1601	16:27	0	572	11535	23544	33814
2/16/15 (7442)	G1215	P03 S1613		-6	668	8351	16661	25651
	G1226	↓	↓	↓	↓	↓	↓	↓
	G1213	P03 S1618		↓	↓	↓	↓	↓
2/16/15 (13001)	G1223	P03 S1619	12:38	-16	567	10573	21592	31813
	G1230	↓	↓	↓	↓	↓	↓	↓
	G1231	↓	↓	↓	↓	↓	↓	↓
	G1224	P03 S1620	↓	↓	↓	↓	↓	↓
2/16/15 (24511)	G1235	P03 S1640	10:19	42	681	8459	16319	26148
	G1240	↓	↓	↓	↓	↓	↓	↓
	G1245	↓	↓	↓	↓	↓	↓	↓
2/14/15 (7443)	G1252	P03 S1647	12:58	-13	576	11625	23862	35254
2/16/15 (13001)	G1247	P03 S1664	9:30	14	512	10240	20432	28863
2/16/15 (7444)	G1254	P03 S1669	14:52	-43	763	11630	24034	35459
2/16/15 (7446)	G1253	P03 S1689	8:58	-27	754	7949	15670	29367
	G1253	P03 S1690	↓	↓	↓	↓	↓	↓
2/16/15 (13001)	G1190	P03 S1686	11:21	51	863	10257	20292	29862
	G1203	↓	↓	↓	↓	↓	↓	↓
	G1204	P03 S1687	↓	↓	↓	↓	↓	↓

MERCURY ANALYSIS LOGBOOK

[illegible]


ISM01.3

LB74791
INSTRUMENT ID: CV1

Linear

 μ Abs.:

40717



A= 0.0000e+000

B= 2.4764e-004

C= -1.1628e-002

Rho= 0.9998794

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/D
0.0	0.000	0.001	0.001	51	0.000	51					1
0.2	0.200	0.202	0.002	863	0.0 %	863					1
2.5	2.500	2.528	0.028	10257	0.0 %	10257					1
5.0	5.000	5.014	0.014	20292	0.0 %	20292					0
7.5	7.500	7.383	-0.117	29862	0.0 %	29862					2
10.0	10.000	10.072	0.072	40717	0.0 %	40717					1

nb
2/11/15

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0		51	-		0 ISM01.3	PPB	2/11/2015 11:21	Std
0.2 S0.2		863	-		0.2 ISM01.3	PPB	2/11/2015 11:23	Std
2.5 S2.5		10257	-		2.5 ISM01.3	PPB	2/11/2015 11:27	Std
5 S5		20292	-		5 ISM01.3	PPB	2/11/2015 11:29	Std
7.5 S7.5		29862	-		7.5 ISM01.3	PPB	2/11/2015 11:31	Std
10 S10		40717	-		10 ISM01.3	PPB	2/11/2015 11:33	Std
ICV31	ICV31	16921	4.1787 -		ISM01.3	PPB	2/11/2015 11:37	SMPL
ICB31	ICB31	-118	-0.0408 -		ISM01.3	PPB	2/11/2015 11:39	SMPL
CCV05	CCV05	20078	4.9605 -		ISM01.3	PPB	2/11/2015 11:41	SMPL
CCB05	CCB05	-123	-0.0421 -		ISM01.3	PPB	2/11/2015 11:43	SMPL
PB81686BL	PBW02	2	-0.0111 -		ISM01.3	PPB	2/11/2015 11:45	SMPL
G1190-13	ME54C7	-33	-0.0198 -		ISM01.3	PPB	2/11/2015 11:47	SMPL
G1190-21	ME54C8	-18	-0.0161 -		ISM01.3	PPB	2/11/2015 11:49	SMPL
G1203-01	ME5D54	-48	-0.0235 -		ISM01.3	PPB	2/11/2015 11:51	SMPL
G1203-04	ME5D62	-43	-0.0223 -		ISM01.3	PPB	2/11/2015 11:53	SMPL
G1203-05	ME5D66	-20	-0.0166 -		ISM01.3	PPB	2/11/2015 11:56	SMPL
G1203-06	ME5D85	-13	-0.0148 -		ISM01.3	PPB	2/11/2015 11:58	SMPL
G1203-07	ME5D86	-30	-0.0191 -		ISM01.3	PPB	2/11/2015 12:00	SMPL
G1203-08	ME5D88	-40	-0.0215 -		ISM01.3	PPB	2/11/2015 12:02	SMPL
G1203-09	ME5D98	-28	-0.0186 -		ISM01.3	PPB	2/11/2015 12:04	SMPL
G1203-10	ME5DA2	-35	-0.0203 -		ISM01.3	PPB	2/11/2015 12:06	SMPL
G1203-11	ME5DA2D	-6	-0.0131 -		ISM01.3	PPB	2/11/2015 12:08	SMPL
G1203-12	ME5DA2S	4321	1.0584 -		ISM01.3	PPB	2/11/2015 12:10	SMPL
G1203-13	ME5DB0	-65	-0.0277 -		ISM01.3	PPB	2/11/2015 12:12	SMPL
PB81687BL	PBW02	-28	-0.0186 -		ISM01.3	PPB	2/11/2015 12:15	SMPL
G1204-01	ME5D77	-60	-0.0265 -		ISM01.3	PPB	2/11/2015 12:17	SMPL
G1204-02	ME5D78	-27	-0.0183 -		ISM01.3	PPB	2/11/2015 12:19	SMPL
G1204-05	ME5D79	-25	-0.0178 -		ISM01.3	PPB	2/11/2015 12:21	SMPL
G1204-06	ME5D84	-33	-0.0198 -		ISM01.3	PPB	2/11/2015 12:23	SMPL
G1204-07	ME5D87	-27	-0.0183 -		ISM01.3	PPB	2/11/2015 12:25	SMPL
CCV06	CCV06	20367	5.0321 -		ISM01.3	PPB	2/11/2015 12:27	SMPL
CCB06	CCB06	-88	-0.0334 -		ISM01.3	PPB	2/11/2015 12:29	SMPL
G1204-08	ME5D89	-29	-0.0188 -		ISM01.3	PPB	2/11/2015 12:32	SMPL
G1204-09	ME5D99	-40	-0.0215 -		ISM01.3	PPB	2/11/2015 12:34	SMPL
G1204-10	ME5DA4	-39	-0.0213 -		ISM01.3	PPB	2/11/2015 12:36	SMPL
G1204-11	ME5DA4D	-40	-0.0215 -		ISM01.3	PPB	2/11/2015 12:38	SMPL
G1204-12	ME5DA4S	4065	0.995 -		ISM01.3	PPB	2/11/2015 12:40	SMPL
G1204-13	ME5DB1	-62	-0.027 -		ISM01.3	PPB	2/11/2015 12:42	SMPL
CCV07	CCV07	20636	5.0987 -		ISM01.3	PPB	2/11/2015 12:44	SMPL
CCB07	CCB07	-98	-0.0359 -		ISM01.3	PPB	2/11/2015 12:46	SMPL