

MERCURY ANALYSIS LOGBOOK

Date	Case Number	Batch Number	Start Time	BLK	STD1	STD2	STD3	STD4
2/3/15 (74901)	G1199	PB S1575	13:53	-14	779	10069	20215	28588
↓								
2/3/15 (74902)	G1213	PB S1599	9:10	22	875	10638	21215	31976
↓								
2/3/15 (74903)	G1214							
2/3/15 (74904)	G1212	PB S1601	16:27	0	572	11835	23544	35834
↓								
2/3/15 (74905)	G1215	PB S1617		-6	608	8351	16661	25651
↓								
2/3/15 (74906)	G1226							
↓								
2/3/15 (74907)	G1217	PB S1618						
↓								
2/3/15 (74908)	G1223	PB S1619	12:34	-16	867	10573	21592	31817
↓								
2/3/15 (74909)	G1230							
↓								
2/3/15 (74910)	G1231							
↓								
2/3/15 (74911)	G1224	PB S1620						
↓								
2/3/15 (74912)	G1235	PB S1640	10:19	42	681	5459	16719	26148
↓								
2/3/15 (74913)	G1240							
↓								
2/3/15 (74914)	G1245							
↓								
2/3/15 (74915)	G1251							
↓								
2/3/15 (74916)	G1252	PB S1647	12:58	-13	876	11629	23862	35724
↓								
2/3/15 (74917)	G1247	PB S1664	9:30	14	512	10240	20432	28963
↓								

QA Control # A3041058

Page 16

MERCURY ANALYSIS LOGBOOK

[illegible]

QA Control # A3041058

Page 17


ISM01.3

LIB 74772
INSTRUMENT ID: CV1

Linear

 μ Abs.:

40742



A= 0.0000e+000

B= 2.4696e-004

C= -5.9174e-003

Rho= 0.9998921

Accept=Accepted

slope
 $y = mx + intercept$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.002	-0.002	14	0.000	14					-
0.2	0.200	0.195	-0.005	812	0.0 %	812					-3
2.5	2.500	2.523	0.023	10240	0.0 %	10240					1
5.0	5.000	5.040	0.040	20432	0.0 %	20432					1
7.5	7.500	7.389	-0.111	29943	0.0 %	29943					-1
10.0	10.000	10.056	0.056	40742	0.0 %	40742					1

MB
2/10/15

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	14	-		0 ISM01.3	PPB	2/10/2015 9:30	Std
	0.2 S0.2	812	-		0.2 ISM01.3	PPB	2/10/2015 9:32	Std
	2.5 S2.5	10240	-		2.5 ISM01.3	PPB	2/10/2015 9:34	Std
	5 S5	20432	-		5 ISM01.3	PPB	2/10/2015 9:36	Std
	7.5 S7.5	29943	-		7.5 ISM01.3	PPB	2/10/2015 9:38	Std
	10 S10	40742	-		10 ISM01.3	PPB	2/10/2015 9:40	Std
ICV28	ICV28	17060	4.2073	-	ISM01.3	PPB	2/10/2015 9:44	SMPL
ICB28	ICB28	-95	-0.0294	-	ISM01.3	PPB	2/10/2015 9:46	SMPL
CCV44	CCV44	20283	5.0032	-	ISM01.3	PPB	2/10/2015 9:48	SMPL
CCB44	CCB44	-122	-0.036	-	ISM01.3	PPB	2/10/2015 9:51	SMPL
PB81664BL	PBW02	23	-0.0002	-	ISM01.3	PPB	2/10/2015 9:53	SMPL
G1247-01	MA4FL1	16	-0.002	-	ISM01.3	PPB	2/10/2015 9:55	SMPL
G1247-02	MA4FL2	-3	-0.0067	-	ISM01.3	PPB	2/10/2015 9:57	SMPL
G1247-03	MA4FL2D	-15	-0.0096	-	ISM01.3	PPB	2/10/2015 9:59	SMPL
G1247-04	MA4FL2S	4022	0.9874	-	ISM01.3	PPB	2/10/2015 10:01	SMPL
G1247-05	MA4FL3	-53	-0.019	-	ISM01.3	PPB	2/10/2015 10:03	SMPL
G1247-06	MA4FL4	10	-0.0034	-	ISM01.3	PPB	2/10/2015 10:05	SMPL
G1247-07	MA4FL5	-1	-0.0062	-	ISM01.3	PPB	2/10/2015 10:07	SMPL
G1247-08	MA4FL6	56	0.0079	-	ISM01.3	PPB	2/10/2015 10:10	SMPL
G1247-09	MA4FL7	18	-0.0015	-	ISM01.3	PPB	2/10/2015 10:12	SMPL
G1247-11	MA4FM4	8535	2.1019	-	ISM01.3	PPB	2/10/2015 10:14	SMPL
G1247-13	MA4FM7	-65	-0.022	-	ISM01.3	PPB	2/10/2015 10:16	SMPL
G1247-14	MA4FM9	246	0.0548	-	ISM01.3	PPB	2/10/2015 10:18	SMPL
G1247-15	MA4FN1	634	0.1507	-	ISM01.3	PPB	2/10/2015 10:20	SMPL
G1247-16	MA4FN3	26	0.0005	-	ISM01.3	PPB	2/10/2015 10:22	SMPL
G1247-17	MA4FN4	101	0.019	-	ISM01.3	PPB	2/10/2015 10:24	SMPL
G1247-18	MA4FN5	-39	-0.0155	-	ISM01.3	PPB	2/10/2015 10:27	SMPL
G1247-19	MA4FN6	-36	-0.0148	-	ISM01.3	PPB	2/10/2015 10:29	SMPL
G1247-20	MA4FN7	51	0.0067	-	ISM01.3	PPB	2/10/2015 10:31	SMPL
G1247-21	MA4FN8	94	0.0173	-	ISM01.3	PPB	2/10/2015 10:33	SMPL
CCV45	CCV45	20083	4.9539	-	ISM01.3	PPB	2/10/2015 10:41	SMPL
CCB45	CCB45	-112	-0.0336	-	ISM01.3	PPB	2/10/2015 10:43	SMPL