

MERCURY ANALYSIS LOGBOOK

[illegible]

MERCURY ANALYSIS LOGBOOK

[illegible]


ISM01.3

LB74893
INSTRUMENT ID: LV1

Linear

 μ Abs.:

47351



A= 0.0000e+000

B= 2.1175e-004

C= -6.7114e-004

Rho= 0.9999848

Accept=Accepted

slope
Y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.004	0.004	20	0.000	20					-
0.2	0.200	0.209	0.009	991	0.0 %	991					4
2.5	2.500	2.513	0.013	11870	0.0 %	11870					1
5.0	5.000	4.964	-0.036	23445	0.0 %	23445					-1
7.5	7.500	7.484	-0.016	35348	0.0 %	35348					0
10.0	10.000	10.026	0.026	47351	0.0 %	47351					0

LB
2/20/15

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0		20	-	0	ISM01.3	PPB	2/20/2015 14:43	Std
0.2 S0.2		991	-	0.2	ISM01.3	PPB	2/20/2015 14:45	Std
2.5 S2.5		11870	-	2.5	ISM01.3	PPB	2/20/2015 14:49	Std
5 S5		23445	-	5	ISM01.3	PPB	2/20/2015 14:51	Std
7.5 S7.5		35348	-	7.5	ISM01.3	PPB	2/20/2015 14:54	Std
10 S10		47351	-	10	ISM01.3	PPB	2/20/2015 14:56	Std
ICV46	ICV46	18323	3.8793 -		ISM01.3	PPB	2/20/2015 15:00	SMPL
ICB46	ICB46	-74	-0.0163 -		ISM01.3	PPB	2/20/2015 15:02	SMPL
CCV48	CCV48	22866	4.8413 -		ISM01.3	PPB	2/20/2015 15:04	SMPL
CCB48	CCB48	-78	-0.0172 -		ISM01.3	PPB	2/20/2015 15:06	SMPL
PB81772BL	PBS02	-45	-0.0102 -		ISM01.3	PPB	2/20/2015 15:08	SMPL
G1261-01	ME54B7	251	0.0525 -		ISM01.3	PPB	2/20/2015 15:10	SMPL
G1261-02	ME54B7D	268	0.0561 -		ISM01.3	PPB	2/20/2015 15:12	SMPL
G1261-03	ME54B7S	11123	2.3547 -		ISM01.3	PPB	2/20/2015 15:14	SMPL
G1261-04	ME54C0	4207	0.8902 -		ISM01.3	PPB	2/20/2015 15:16	SMPL
G1261-05	ME54C1	46	0.0091 -		ISM01.3	PPB	2/20/2015 15:19	SMPL
G1261-06	ME54C2	49	0.0097 -		ISM01.3	PPB	2/20/2015 15:21	SMPL
CCV49	CCV49	22753	4.8174 -		ISM01.3	PPB	2/20/2015 15:23	SMPL
CCB49	CCB49	-82	-0.018 -		ISM01.3	PPB	2/20/2015 15:25	SMPL