

MERCURY ANALYSIS LOGBOOK

Date	Case Number	Batch Number	Start Time	BLK	STD1	STD2	STD3	STD4
2/24/15	G1329	P1381535	14:34	65	1015	12135	24045	3594
2/25/15	G1350	P1381807	11:58	73	1017	12023	24034	36203
	G1357	P1381508						
	G1330	P1381870						
2/26/15	G1324	P1381584	15:39	-5	950	11818	24074	33748
2/26/15	G1350	P1381571	15:51	-223	816	10999	22246	33309
	G1362							
	G1363							
2/27/15	G1330	P1381555	11:29	-10	734	8764	17957	26237
	G1362							
	G1385	P1381511						
2/27/15	G1330	P1381501	13:14	59	934	11088	23259	32392
	G1348							
	G1374							

MERCURY ANALYSIS LOGBOOK

STD 5	End Time	Correlation Coefficient	Stannous Chloride Prep Log #	Comment	Analyst	Supervisor Signature
47768	15:53	0.9999922	M1726123	ICV13-30 CEV13-16-18	MS	MS
47813	14:48	0.9999510	M1726154	ICV13-01 CEV13-19-23	MS	MS
					MS	MS
46007	16:57	0.9995071	M1726139	ICV13-03 CEV13-24-28	MS	MS
46049	16:57	0.9998014	M1726184	ICV13-02 CEV13-28-30	MS	MS
					MS	MS
34588	12:49	0.9999009	M1726226	ICV13-03 CEV13-31-34	MS	MS
					MS	MS
44366	14:18	0.9994166	M1726226	ICV13-04 CEV13-35-36	MS	MS
					MS	MS
					MS	MS

LB 74 967


INSTRUMENT ID: CV 1

ISM01.3

Linear

 μ Abs.:

4436E



A= 0.0000e+000

B= 2.2669e-004 slope

C= -3.5173e-002 y-intercept

Rho= 0.9994166

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.022	-0.022	59	0.000	59					-
0.2	0.200	0.177	-0.023	934	0.0 %	934					-12
2.5	2.500	2.478	-0.022	11088	0.0 %	11088					-1
5.0	5.000	5.237	0.237	23259	0.0 %	23259					5
7.5	7.500	7.308	-0.192	32392	0.0 %	32392					-3
10.0	10.000	10.022	0.022	44366	0.0 %	44366					0

MB

2/27/15

LB74967
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	59	-		0 ISM01.3	PPB	2/27/2015 13:14	Std
	0.2 S0.2	934	-		0.2 ISM01.3	PPB	2/27/2015 13:17	Std
	2.5 S2.5	11088	-		2.5 ISM01.3	PPB	2/27/2015 13:19	Std
	5 S5	23259	-		5 ISM01.3	PPB	2/27/2015 13:23	Std
	7.5 S7.5	32392	-		7.5 ISM01.3	PPB	2/27/2015 13:25	Std
	10 S10	44366	-		10 ISM01.3	PPB	2/27/2015 13:30	Std
ICV04	ICV04	16089	3.612 -		ISM01.3	PPB	2/27/2015 13:32	SMPL
ICB04	ICB04	-51	-0.0467 -		ISM01.3	PPB	2/27/2015 13:35	SMPL
CCV35	CCV35	19551	4.3968 -		ISM01.3	PPB	2/27/2015 13:37	SMPL
CCB35	CCB35	-56	-0.0479 -		ISM01.3	PPB	2/27/2015 13:39	SMPL
PB81901BL	PBW02	12	-0.0325 -		ISM01.3	PPB	2/27/2015 13:41	SMPL
G1330-20	MJG3D8	-29	-0.0417 -		ISM01.3	PPB	2/27/2015 13:43	SMPL
G1330-21	MJG3D9	-47	-0.0458 -		ISM01.3	PPB	2/27/2015 13:45	SMPL
G1330-22	MJG3E5	470	0.0714 -		ISM01.3	PPB	2/27/2015 13:47	SMPL
G1348-01	MBCDJ2	-96	-0.0569 -		ISM01.3	PPB	2/27/2015 13:49	SMPL
G1374-01	MJG3S8	-32	-0.0424 -		ISM01.3	PPB	2/27/2015 13:52	SMPL
G1374-02	MJG3T7	-76	-0.0524 -		ISM01.3	PPB	2/27/2015 13:54	SMPL
G1374-03	MJG3T8	-52	-0.047 -		ISM01.3	PPB	2/27/2015 13:56	SMPL
G1374-04	MJG3W0	-62	-0.0492 -		ISM01.3	PPB	2/27/2015 13:58	SMPL
G1374-05	MJG3W0D	-66	-0.0501 -		ISM01.3	PPB	2/27/2015 14:00	SMPL
G1374-06	MJG3W0S	3996	0.8707 -		ISM01.3	PPB	2/27/2015 14:06	SMPL
G1374-07	MJG3W1	-43	-0.0449 -		ISM01.3	PPB	2/27/2015 14:08	SMPL
G1374-08	MJG3W2	-39	-0.044 -		ISM01.3	PPB	2/27/2015 14:10	SMPL
G1374-09	MJG3W3	-46	-0.0456 -		ISM01.3	PPB	2/27/2015 14:12	SMPL
G1374-10	MJG3W8	440	0.0646 -		ISM01.3	PPB	2/27/2015 14:14	SMPL
CCV36	CCV36	19607	4.4095 -		ISM01.3	PPB	2/27/2015 14:16	SMPL
CCB36	CCB36	-94	-0.0565 -		ISM01.3	PPB	2/27/2015 14:18	SMPL