

MERCURY ANALYSIS LOGBOOK

[illegible]

MERCURY ANALYSIS LOGBOOK

[illegible]


ISM01.3

LB 75062
INSTRUMENT ID: CV1

Linear

 μ Abs.:

48027



A= 0.0000e+000

B= 2.0796e-004

C= -2.4032e-002

Rho= 0.9999790

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.010	-0.010	66	0.000	66					-
0.2	0.200	0.184	-0.016	999	0.0 %	999					-8
2.5	2.500	2.518	0.018	12222	0.0 %	12222					1
5.0	5.000	5.008	0.008	24198	0.0 %	24198					0
7.5	7.500	7.537	0.037	36358	0.0 %	36358					0
10.0	10.000	9.964	-0.036	48027	0.0 %	48027					0

NB
3/10/15

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0		66	-		0 ISM01.3	PPB	3/10/2015 15:08	Std
0.2 S0.2		999	-		0.2 ISM01.3	PPB	3/10/2015 15:10	Std
2.5 S2.5		12222	-		2.5 ISM01.3	PPB	3/10/2015 15:12	Std
5 S5		24198	-		5 ISM01.3	PPB	3/10/2015 15:14	Std
7.5 S7.5		36358	-		7.5 ISM01.3	PPB	3/10/2015 15:17	Std
10 S10		48027	-		10 ISM01.3	PPB	3/10/2015 15:19	Std
ICV18	ICV18	18897	3.9058 -		ISM01.3	PPB	3/10/2015 15:23	SMPL
ICB18	ICB18	-31	-0.0305 -		ISM01.3	PPB	3/10/2015 15:25	SMPL
CCV39	CCV39	23557	4.8749 -		ISM01.3	PPB	3/10/2015 15:27	SMPL
CCB39	CCB39	-66	-0.0378 -		ISM01.3	PPB	3/10/2015 15:29	SMPL
PB82021BL	PBS02	41	-0.0155 -		ISM01.3	PPB	3/10/2015 15:31	SMPL
G1446-01	MECD01	660	0.1132 -		ISM01.3	PPB	3/10/2015 15:33	SMPL
G1446-02	MECD02	180	0.0134 -		ISM01.3	PPB	3/10/2015 15:35	SMPL
G1446-03	MECD03	523	0.0847 -		ISM01.3	PPB	3/10/2015 15:37	SMPL
G1446-04	MECD04	252	0.0284 -		ISM01.3	PPB	3/10/2015 15:40	SMPL
G1446-05	MECD05	312	0.0409 -		ISM01.3	PPB	3/10/2015 15:42	SMPL
G1446-06	MECD06	828	0.1482 -		ISM01.3	PPB	3/10/2015 15:44	SMPL
G1446-07	MECD07	833	0.1492 -		ISM01.3	PPB	3/10/2015 15:46	SMPL
G1446-08	MECD08	1939	0.3792 -		ISM01.3	PPB	3/10/2015 15:48	SMPL
G1446-09	MECD09	681	0.1176 -		ISM01.3	PPB	3/10/2015 15:50	SMPL
G1446-10	MECD10	366	0.0521 -		ISM01.3	PPB	3/10/2015 15:52	SMPL
G1446-11	MECD11	448	0.0691 -		ISM01.3	PPB	3/10/2015 15:54	SMPL
G1446-12	MECD12	752	0.1324 -		ISM01.3	PPB	3/10/2015 15:57	SMPL
G1446-13	MECD12D	732	0.1282 -		ISM01.3	PPB	3/10/2015 15:59	SMPL
G1446-14	MECD12S	12205	2.5141 -		ISM01.3	PPB	3/10/2015 16:01	SMPL
G1446-15	MECD13	623	0.1055 -		ISM01.3	PPB	3/10/2015 16:03	SMPL
CCV40	CCV40	24013	4.9697 -		ISM01.3	PPB	3/10/2015 16:05	SMPL
CCB40	CCB40	-97	-0.0442 -		ISM01.3	PPB	3/10/2015 16:07	SMPL