

MERCURY ANALYSIS LOGBOOK

STD 5	End Time	Correlation Coefficient	Stannous Chloride Prep log #	Comment	Analyst	Supervisor Signature
35433	12:02	0.999359	1127150	2013-25 2013-28-31	MB	MB
46520	14:25	0.999373	1127150	2013-26 2013-32-35	MB	MB
42271	15:34	0.999494	1127150	2013-27 2013-36-37	MB	MB
41590	16:45	0.999610	1127150	2013-38-39	MB	MB

MERCURY ANALYSIS LOGBOOK

Date	Case Number	Batch Number	Start Time	BLK	STD1	STD2	STD3	STD4
4/16/15 (7470)	G1905	PB82792	10:34	13	923	5688	17698	26161
4/16/15 (7471)	G1902	PB82793						
4/16/15 (7472)	G1744	PB82794	12:37	121	966	11567	22544	33466
4/16/15 (7473)	G1705							
4/16/15 (7474)	G1891							
4/16/15 (7475)	G1902							
4/16/15 (7476)	G1905							
4/16/15 (7477)	G1915							
4/16/15 (7478)	G1834	PB82811	15:00	80	927	10642	21140	31691
4/16/15 (7479)	G1887	PB82810	15:58	88	917	10407	20992	31049


ISM01.3

LB 75664
INSTRUMENT ID : CV2

Linear

 μ Abs.:

41890



A= 0.0000e+000

B= 2.3993e-004

C= -1.2511e-002

Rho= 0.9999610

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.0	0.000	0.009	0.009	88	0.000	88					4
0.2	0.200	0.208	0.008	917	0.0 %	917					-1
2.5	2.500	2.484	-0.016	10407	0.0 %	10407					0
5.0	5.000	5.024	0.024	20992	0.0 %	20992					-1
7.5	7.500	7.437	-0.063	31049	0.0 %	31049					0
10.0	10.000	10.038	0.038	41890	0.0 %	41890					

NB
4116115

LB75664
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0		88	-	0	ISM01.3	PPB	4/16/2015 15:58	Std
0.2 S0.2		917	-	0.2	ISM01.3	PPB	4/16/2015 16:00	Std
2.5 S2.5		10407	-	2.5	ISM01.3	PPB	4/16/2015 16:02	Std
5 S5		20992	-	5	ISM01.3	PPB	4/16/2015 16:04	Std
7.5 S7.5		31049	-	7.5	ISM01.3	PPB	4/16/2015 16:06	Std
10 S10		41890	-	10	ISM01.3	PPB	4/16/2015 16:09	Std
ICV27	ICV27	16567	3.9624 -		ISM01.3	PPB	4/16/2015 16:11	SMPL
ICB27	ICB27	27	-0.006 -		ISM01.3	PPB	4/16/2015 16:14	SMPL
CCV38	CCV38	20996	5.0251 -		ISM01.3	PPB	4/16/2015 16:16	SMPL
CCB38	CCB38	7	-0.0108 -		ISM01.3	PPB	4/16/2015 16:18	SMPL
PB82810BL	PBW06	117	0.0156 -		ISM01.3	PPB	4/16/2015 16:20	SMPL
G1867-01	MCOAD8	62	0.0024 -		ISM01.3	PPB	4/16/2015 16:22	SMPL
G1867-02	MCOAE0	32	-0.0048 -		ISM01.3	PPB	4/16/2015 16:24	SMPL
G1867-03	MCOAE1	34	-0.0044 -		ISM01.3	PPB	4/16/2015 16:26	SMPL
G1867-04	MCOAE1D	12	-0.0096 -		ISM01.3	PPB	4/16/2015 16:28	SMPL
G1867-05	MCOAE1S	4206	0.9966 -		ISM01.3	PPB	4/16/2015 16:30	SMPL
G1867-06	MCOAE2	-24	-0.0183 -		ISM01.3	PPB	4/16/2015 16:33	SMPL
G1867-07	MCOAE5	-14	-0.0159 -		ISM01.3	PPB	4/16/2015 16:35	SMPL
G1867-08	MCOAE5D	-2	-0.013 -		ISM01.3	PPB	4/16/2015 16:37	SMPL
G1867-09	MCOAE5S	4182	0.9909 -		ISM01.3	PPB	4/16/2015 16:41	SMPL
G1867-10	MCOAE6	-12	-0.0154 -		ISM01.3	PPB	4/16/2015 16:43	SMPL
CCV39	CCV39	20410	4.8845 -		ISM01.3	PPB	4/16/2015 16:46	SMPL
CCB39	CCB39	-39	-0.0219 -		ISM01.3	PPB	4/16/2015 16:48	SMPL