


ISM01.3

LB 75928
INSTRUMENT ID: CV1

Linear

 μ Abs.:

41209



A= 0.0000e+000

B= 2.4183e-004

C= -5.9885e-003

Rho= 0.9999728

Accept= Accepted

slope
y = m/x

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.008	-0.008	-7	0.000	-7					-
0.2	0.200	0.195	-0.005	832	0.0 %	832					-3
2.5	2.500	2.484	-0.016	10295	0.0 %	10295					-1
5.0	5.000	5.031	0.031	20829	0.0 %	20829					1
7.5	7.500	7.538	0.038	31198	0.0 %	31198					1
10.0	10.000	9.959	-0.041	41209	0.0 %	41209					0

LB
5/6/15

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-7	-		0 ISM01.3	PPB	5/6/2015 16:11	Std
	0.2 S0.2	832	-		0.2 ISM01.3	PPB	5/6/2015 16:13	Std
	2.5 S2.5	10295	-		2.5 ISM01.3	PPB	5/6/2015 16:15	Std
	5 S5	20829	-		5 ISM01.3	PPB	5/6/2015 16:17	Std
	7.5 S7.5	31198	-		7.5 ISM01.3	PPB	5/6/2015 16:19	Std
	10 S10	41209	-		10 ISM01.3	PPB	5/6/2015 16:21	Std
ICV09	ICV09	16739	4.0419 -		ISM01.3	PPB	5/6/2015 16:24	SMPL
ICB09	ICB09	-29	-0.013 -		ISM01.3	PPB	5/6/2015 16:26	SMPL
CCV33	CCV33	20948	5.0598 -		ISM01.3	PPB	5/6/2015 16:28	SMPL
CCB33	CCB33	-26	-0.0123 -		ISM01.3	PPB	5/6/2015 16:30	SMPL
PB83212BL	PBW06	58	0.008 -		ISM01.3	PPB	5/6/2015 16:32	SMPL
G2105-01	MBOAA2	639	0.1485 -		ISM01.3	PPB	5/6/2015 16:34	SMPL
G2105-02	MBOAA3	-222	-0.0597 -		ISM01.3	PPB	5/6/2015 16:37	SMPL
G2105-03	MBOAA3D	-68	-0.0224 -		ISM01.3	PPB	5/6/2015 16:39	SMPL
G2105-04	MBOAA3S	3953	0.9499 -		ISM01.3	PPB	5/6/2015 16:41	SMPL
G2105-05	MBOAA4	-104	-0.0311 -		ISM01.3	PPB	5/6/2015 16:43	SMPL
G2105-06	MBOAA5	198	0.0419 -		ISM01.3	PPB	5/6/2015 16:45	SMPL
G2105-07	MBOAA6	-168	-0.0466 -		ISM01.3	PPB	5/6/2015 16:47	SMPL
G2105-08	MBOAA7	-55	-0.0193 -		ISM01.3	PPB	5/6/2015 16:49	SMPL
G2105-09	MBOAA8	-75	-0.0241 -		ISM01.3	PPB	5/6/2015 16:51	SMPL
G2105-10	MBOAA9	-46	-0.0171 -		ISM01.3	PPB	5/6/2015 16:54	SMPL
G2105-11	MBOAB0	509	0.1171 -		ISM01.3	PPB	5/6/2015 16:56	SMPL
G2105-12	MBOAB1	-138	-0.0394 -		ISM01.3	PPB	5/6/2015 16:58	SMPL
G2105-13	MBOAB2	-73	-0.0236 -		ISM01.3	PPB	5/6/2015 17:00	SMPL
G2105-14	MBOAB3	-71	-0.0232 -		ISM01.3	PPB	5/6/2015 17:02	SMPL
G2105-15	MBOAB4	-109	-0.0323 -		ISM01.3	PPB	5/6/2015 17:04	SMPL
G2105-16	MBOAB5	-93	-0.0285 -		ISM01.3	PPB	5/6/2015 17:06	SMPL
G2105-17	MBOAB6	-131	-0.0377 -		ISM01.3	PPB	5/6/2015 17:08	SMPL
G2105-18	MBOAB7	-100	-0.0302 -		ISM01.3	PPB	5/6/2015 17:10	SMPL
G2105-19	MBOAB8	-62	-0.021 -		ISM01.3	PPB	5/6/2015 17:13	SMPL
CCV34	CCV34	20075	4.8487 -		ISM01.3	PPB	5/6/2015 17:15	SMPL
CCB34	CCB34	-126	-0.0365 -		ISM01.3	PPB	5/6/2015 17:17	SMPL
G2105-20	MBOAD7	141	0.0281 -		ISM01.3	PPB	5/6/2015 17:19	SMPL
G2105-21	MBOAD8	-94	-0.0287 -		ISM01.3	PPB	5/6/2015 17:21	SMPL
G2105-22	MBOAD9	225	0.0484 -		ISM01.3	PPB	5/6/2015 17:23	SMPL
PB83213BL	PBW06	-73	-0.0236 -		ISM01.3	PPB	5/6/2015 17:25	SMPL
G2125-11	MBOAC7	-45	-0.0169 -		ISM01.3	PPB	5/6/2015 17:27	SMPL
G2125-12	MBOAC7D	-35	-0.0145 -		ISM01.3	PPB	5/6/2015 17:30	SMPL
G2125-13	MBOAC7S	3858	0.927 -		ISM01.3	PPB	5/6/2015 17:32	SMPL
G2125-01	MBOAC1	-116	-0.034 -		ISM01.3	PPB	5/6/2015 17:34	SMPL
G2125-02	MBOAC3	-68	-0.0224 -		ISM01.3	PPB	5/6/2015 17:36	SMPL
G2125-03	MBOAC5	45	0.0049 -		ISM01.3	PPB	5/6/2015 17:38	SMPL
G2125-04	MBOAC6	-90	-0.0278 -		ISM01.3	PPB	5/6/2015 17:40	SMPL
G2125-05	MBOAE4	24	-0.0002 -		ISM01.3	PPB	5/6/2015 17:42	SMPL
G2125-06	MBOAE5	278	0.0612 -		ISM01.3	PPB	5/6/2015 17:44	SMPL
G2125-07	MBOAE6	-68	-0.0224 -		ISM01.3	PPB	5/6/2015 17:46	SMPL

LB75928
INSTRUMENT ID : CV1

G2125-08	MB0AE7	16	-0.0021 -	ISM01.3	PPB	5/6/2015 17:49	SMPL
G2125-09	MB0AE8	-61	-0.0207 -	ISM01.3	PPB	5/6/2015 17:51	SMPL
G2125-10	MB0AF0	-59	-0.0203 -	ISM01.3	PPB	5/6/2015 17:53	SMPL
G2125-14	MB0AC8	64	0.0095 -	ISM01.3	PPB	5/6/2015 17:55	SMPL
G2125-15	MB0AC9	-49	-0.0178 -	ISM01.3	PPB	5/6/2015 17:57	SMPL
G2125-16	MB0AD0	-54	-0.019 -	ISM01.3	PPB	5/6/2015 17:59	SMPL
CCV35	CCV35	20560	4.9659 -	ISM01.3	PPB	5/6/2015 18:01	SMPL
CCB35	CCB35	-147	-0.0415 -	ISM01.3	PPB	5/6/2015 18:03	SMPL
G2125-17	MB0AD1	159	0.0325 -	ISM01.3	PPB	5/6/2015 18:06	SMPL
G2125-18	MB0AD2	9	-0.0038 -	ISM01.3	PPB	5/6/2015 18:08	SMPL
G2125-19	MB0AD3	7	-0.0043 -	ISM01.3	PPB	5/6/2015 18:10	SMPL
G2125-20	MB0AD4	-79	-0.0251 -	ISM01.3	PPB	5/6/2015 18:12	SMPL
G2125-21	MB0AD5	17	-0.0019 -	ISM01.3	PPB	5/6/2015 18:14	SMPL
G2125-22	MB0AD6	-55	-0.0193 -	ISM01.3	PPB	5/6/2015 18:16	SMPL
CCV36	CCV36	20502	4.9519 -	ISM01.3	PPB	5/6/2015 18:18	SMPL
CCB36	CCB36	-208	-0.0563 -	ISM01.3	PPB	5/6/2015 18:20	SMPL