

MERCURY ANALYSIS LOGBOOK

[illegible]

OA Control # A3041058

Page 44

MERCURY ANALYSIS LOGBOOK

[illegible]

QA Control # A3041058

LB76029 CV1

Page 45


ISM01.3

LB76029
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42357



A= 0.0000e+000

B= 2.3635e-004

C= -2.6045e-002

Rho= 0.9999944

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.006	0.006	137	0.000	137					-
0.2	0.200	0.180	-0.020	873	0.0 %	873					-10
2.5	2.500	2.508	0.008	10720	0.0 %	10720					0
5.0	5.000	5.009	0.009	21302	0.0 %	21302					0
7.5	7.500	7.512	0.012	31892	0.0 %	31892					0
10.0	10.000	9.985	-0.015	42357	0.0 %	42357					0

NB
5/13/15

LB76029
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0		137	-		0 ISM01.3	PPB	5/13/2015 12:42	Std
0.2 S0.2		873	-		0.2 ISM01.3	PPB	5/13/2015 12:44	Std
2.5 S2.5		10720	-		2.5 ISM01.3	PPB	5/13/2015 12:46	Std
5 S5		21302	-		5 ISM01.3	PPB	5/13/2015 12:48	Std
7.5 S7.5		31892	-		7.5 ISM01.3	PPB	5/13/2015 12:50	Std
10 S10		42357	-		10 ISM01.3	PPB	5/13/2015 12:53	Std
ICV20	ICV20	17621	4.1387	-	ISM01.3	PPB	5/13/2015 12:55	SMPL
ICB20	ICB20	52	-0.0138	-	ISM01.3	PPB	5/13/2015 12:57	SMPL
CCV25	CCV25	20177	4.7429	-	ISM01.3	PPB	5/13/2015 13:00	SMPL
CCB25	CCB25	81	-0.0069	-	ISM01.3	PPB	5/13/2015 13:02	SMPL
PB83348BL	PBW06	103	-0.0017	-	ISM01.3	PPB	5/13/2015 13:04	SMPL
G2126-01	MB0AC2	614	0.1191	-	ISM01.3	PPB	5/13/2015 13:06	SMPL
G2126-02	MB0AC4	-67	-0.0419	-	ISM01.3	PPB	5/13/2015 13:08	SMPL
G2126-03	MB0AC4D	-31	-0.0334	-	ISM01.3	PPB	5/13/2015 13:10	SMPL
G2126-04	MB0AC4S	3980	0.9146	-	ISM01.3	PPB	5/13/2015 13:12	SMPL
G2126-05	MB0AE0	242	0.0312	-	ISM01.3	PPB	5/13/2015 13:14	SMPL
G2126-06	MB0AE1	36	-0.0175	-	ISM01.3	PPB	5/13/2015 13:17	SMPL
G2126-07	MB0AE2	0	-0.026	-	ISM01.3	PPB	5/13/2015 13:19	SMPL
G2126-08	MB0AE3	2886	0.6561	-	ISM01.3	PPB	5/13/2015 13:21	SMPL
G2126-09	MB0AE9	86	-0.0057	-	ISM01.3	PPB	5/13/2015 13:23	SMPL
G2126-10	MB0AF1	172	0.0146	-	ISM01.3	PPB	5/13/2015 13:25	SMPL
G2126-11	MB0AF4	-5	-0.0272	-	ISM01.3	PPB	5/13/2015 13:27	SMPL
G2126-12	MB0AF5	-29	-0.0329	-	ISM01.3	PPB	5/13/2015 13:29	SMPL
G2126-13	MB0AF6	-24	-0.0317	-	ISM01.3	PPB	5/13/2015 13:31	SMPL
G2126-14	MB0AF7	54	-0.0133	-	ISM01.3	PPB	5/13/2015 13:33	SMPL
G2126-15	MB0AF2	-114	-0.053	-	ISM01.3	PPB	5/13/2015 13:36	SMPL
G2126-16	MB0AF3	-82	-0.0454	-	ISM01.3	PPB	5/13/2015 13:38	SMPL
G2126-17	MB0AF8	49	-0.0145	-	ISM01.3	PPB	5/13/2015 13:40	SMPL
G2126-18	MB0AF9	10	-0.0237	-	ISM01.3	PPB	5/13/2015 13:42	SMPL
G2126-19	MB0AG0	-83	-0.0457	-	ISM01.3	PPB	5/13/2015 13:44	SMPL
CCV26	CCV26	19735	4.6384	-	ISM01.3	PPB	5/13/2015 13:46	SMPL
CCB26	CCB26	-88	-0.0468	-	ISM01.3	PPB	5/13/2015 13:48	SMPL
G2126-20	MB0AG1	106	-0.001	-	ISM01.3	PPB	5/13/2015 13:50	SMPL
G2126-21	MB0AG2	-105	-0.0509	-	ISM01.3	PPB	5/13/2015 13:53	SMPL
G2126-22	MB0AG3	-66	-0.0416	-	ISM01.3	PPB	5/13/2015 13:55	SMPL
PB83354BL	PBW06	5	-0.0249	-	ISM01.3	PPB	5/13/2015 13:57	SMPL
G2162-01	ME5GE4	-68	-0.0421	-	ISM01.3	PPB	5/13/2015 13:59	SMPL
G2162-02	ME5GE4D	-89	-0.0471	-	ISM01.3	PPB	5/13/2015 14:01	SMPL
G2162-03	ME5GE4S	4036	0.9279	-	ISM01.3	PPB	5/13/2015 14:03	SMPL
G2162-04	ME5GJ4	-46	-0.0369	-	ISM01.3	PPB	5/13/2015 14:05	SMPL
G2162-05	ME5GJ6	-31	-0.0334	-	ISM01.3	PPB	5/13/2015 14:07	SMPL
G2162-06	ME5GL9	-27	-0.0324	-	ISM01.3	PPB	5/13/2015 14:09	SMPL
G2162-07	ME5GH2	5	-0.0249	-	ISM01.3	PPB	5/13/2015 14:12	SMPL
G2162-08	ME5GH6	-2	-0.0265	-	ISM01.3	PPB	5/13/2015 14:14	SMPL
G2162-09	ME5GJ8	17	-0.022	-	ISM01.3	PPB	5/13/2015 14:16	SMPL
G2162-10	ME5GK0	24	-0.0204	-	ISM01.3	PPB	5/13/2015 14:18	SMPL

LB76029
INSTRUMENT ID : CV1

G2162-11	ME5GK1	63	-0.0112 -	ISM01.3	PPB	5/13/2015 14:20	SMPL
G2162-12	ME5GH4	44	-0.0156 -	ISM01.3	PPB	5/13/2015 14:22	SMPL
G2162-13	ME5GH8	95	-0.0036 -	ISM01.3	PPB	5/13/2015 14:24	SMPL
G2162-14	ME5GL3	-48	-0.0374 -	ISM01.3	PPB	5/13/2015 14:26	SMPL
G2162-15	ME5GL4	-44	-0.0364 -	ISM01.3	PPB	5/13/2015 14:29	SMPL
G2162-16	ME5GK4	-15	-0.0296 -	ISM01.3	PPB	5/13/2015 14:31	SMPL
CCV27	CCV27	19906	4.6788 -	ISM01.3	PPB	5/13/2015 14:33	SMPL
CCB27	CCB27	-54	-0.0388 -	ISM01.3	PPB	5/13/2015 14:35	SMPL
G2162-17	ME5GK7	10	-0.0237 -	ISM01.3	PPB	5/13/2015 14:37	SMPL
G2162-18	ME5GL0	11	-0.0234 -	ISM01.3	PPB	5/13/2015 14:39	SMPL
PB83355BL	PBW06	18	-0.0218 -	ISM01.3	PPB	5/13/2015 14:41	SMPL
G2167-01	ME5GM2	-4	-0.027 -	ISM01.3	PPB	5/13/2015 14:43	SMPL
G2167-02	ME5GM2D	-29	-0.0329 -	ISM01.3	PPB	5/13/2015 14:46	SMPL
G2167-03	ME5GM2S	4085	0.9395 -	ISM01.3	PPB	5/13/2015 14:48	SMPL
G2167-04	ME5GG0	188	0.0184 -	ISM01.3	PPB	5/13/2015 14:50	SMPL
G2167-05	ME5GG3	-89	-0.0471 -	ISM01.3	PPB	5/13/2015 14:52	SMPL
G2167-06	ME5GH9	-20	-0.0308 -	ISM01.3	PPB	5/13/2015 14:54	SMPL
G2167-07	ME5GF0	10	-0.0237 -	ISM01.3	PPB	5/13/2015 14:56	SMPL
G2167-08	ME5GG5	3	-0.0253 -	ISM01.3	PPB	5/13/2015 14:58	SMPL
G2167-09	ME5GJ1	-18	-0.0303 -	ISM01.3	PPB	5/13/2015 15:00	SMPL
G2167-10	ME5GJ2	-55	-0.039 -	ISM01.3	PPB	5/13/2015 15:03	SMPL
G2167-11	ME5GJ5	-58	-0.0398 -	ISM01.3	PPB	5/13/2015 15:05	SMPL
G2167-12	ME5GH1	-22	-0.0312 -	ISM01.3	PPB	5/13/2015 15:07	SMPL
G2167-13	ME5GH5	-9	-0.0282 -	ISM01.3	PPB	5/13/2015 15:09	SMPL
G2167-14	ME5GJ7	-121	-0.0546 -	ISM01.3	PPB	5/13/2015 15:11	SMPL
G2167-15	ME5GJ9	-104	-0.0506 -	ISM01.3	PPB	5/13/2015 15:13	SMPL
G2167-16	ME5GK2	-80	-0.045 -	ISM01.3	PPB	5/13/2015 15:15	SMPL
G2167-17	ME5GH3	-76	-0.044 -	ISM01.3	PPB	5/13/2015 15:17	SMPL
CCV28	CCV28	19929	4.6842 -	ISM01.3	PPB	5/13/2015 15:19	SMPL
CCB28	CCB28	-92	-0.0478 -	ISM01.3	PPB	5/13/2015 15:22	SMPL
G2167-18	ME5GH7	-84	-0.0459 -	ISM01.3	PPB	5/13/2015 15:24	SMPL
G2167-19	ME5GL1	-150	-0.0615 -	ISM01.3	PPB	5/13/2015 15:26	SMPL
G2167-20	ME5GL2	-97	-0.049 -	ISM01.3	PPB	5/13/2015 15:28	SMPL
G2167-21	ME5GL8	-98	-0.0492 -	ISM01.3	PPB	5/13/2015 15:30	SMPL
G2167-22	ME5GK3	-85	-0.0461 -	ISM01.3	PPB	5/13/2015 15:32	SMPL
PB83356BL	PBW06	11	-0.0234 -	ISM01.3	PPB	5/13/2015 15:34	SMPL
G2213-01	MC0AF5	-25	-0.032 -	ISM01.3	PPB	5/13/2015 15:37	SMPL
G2213-02	MC0AF6	-32	-0.0336 -	ISM01.3	PPB	5/13/2015 15:39	SMPL
G2213-03	MC0AF6D	-16	-0.0298 -	ISM01.3	PPB	5/13/2015 15:41	SMPL
G2213-04	MC0AF6S	4140	0.9525 -	ISM01.3	PPB	5/13/2015 15:43	SMPL
G2213-05	MC0AF7	-46	-0.0369 -	ISM01.3	PPB	5/13/2015 15:45	SMPL
G2213-06	MC0AF8	-69	-0.0424 -	ISM01.3	PPB	5/13/2015 15:47	SMPL
G2213-07	MC0AF9	-51	-0.0381 -	ISM01.3	PPB	5/13/2015 15:49	SMPL
G2213-08	MC0AGO	-78	-0.0445 -	ISM01.3	PPB	5/13/2015 15:51	SMPL
G2213-09	MC0AG1	-45	-0.0367 -	ISM01.3	PPB	5/13/2015 15:54	SMPL
PB83357BL	PBW06	-56	-0.0393 -	ISM01.3	PPB	5/13/2015 15:56	SMPL
G2191-01	MH0AH6	-90	-0.0473 -	ISM01.3	PPB	5/13/2015 15:58	SMPL

INSTRUMENT ID : CV1

G2191-02	MH0AH7	-82	-0.0454 -	ISM01.3	PPB	5/13/2015 16:00	SMPL
G2191-03	MH0AH8	-74	-0.0435 -	ISM01.3	PPB	5/13/2015 16:02	SMPL
G2191-04	MH0AH9	-98	-0.0492 -	ISM01.3	PPB	5/13/2015 16:04	SMPL
CCV29	CCV29	19945	4.688 -	ISM01.3	PPB	5/13/2015 16:06	SMPL
CCB29	CCB29	-18	-0.0303 -	ISM01.3	PPB	5/13/2015 16:08	SMPL
G2191-05	MH0AJ0	-72	-0.0431 -	ISM01.3	PPB	5/13/2015 16:11	SMPL
G2191-06	MH0AJ0D	-48	-0.0374 -	ISM01.3	PPB	5/13/2015 16:13	SMPL
G2191-07	MH0AJ0S	4153	0.9555 -	ISM01.3	PPB	5/13/2015 16:18	SMPL
G2191-08	MH0AJ1	-97	-0.049 -	ISM01.3	PPB	5/13/2015 16:20	SMPL
G2191-09	MH0AJ4	-80	-0.045 -	ISM01.3	PPB	5/13/2015 16:22	SMPL
PB83358BL	PBW06	-31	-0.0334 -	ISM01.3	PPB	5/13/2015 16:25	SMPL
G2192-01	MH0AJ2	-63	-0.0409 -	ISM01.3	PPB	5/13/2015 16:27	SMPL
G2192-02	MH0AJ3	-82	-0.0454 -	ISM01.3	PPB	5/13/2015 16:29	SMPL
G2192-03	MH0AJ5	-90	-0.0473 -	ISM01.3	PPB	5/13/2015 16:31	SMPL
G2192-04	MH0AJ5D	-84	-0.0459 -	ISM01.3	PPB	5/13/2015 16:33	SMPL
G2192-05	MH0AJ5S	3741	0.8582 -	ISM01.3	PPB	5/13/2015 16:38	SMPL
CCV30	CCV30	20137	4.7334 -	ISM01.3	PPB	5/13/2015 16:40	SMPL
CCB30	CCB30	10	-0.0237 -	ISM01.3	PPB	5/13/2015 16:42	SMPL