

MERCURY ANALYSIS LOGBOOK

[illegible]

QA Control # A3041058

Page 48

MERCURY ANALYSIS LOGBOOK[illegible]

QA Control # A3041058

Page 49

LB 76128 CVI.


ISM01.3

LB76128
INSTRUMENT IDLEV1

Linear

 μ Abs.:

42593



A= 0.0000e+000

B= 2.3353e-004

C= 4.4142e-002

Rho= 0.9994778

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.052	0.052	34	0.000	34					13
0.2	0.200	0.226	0.026	777	0.0 %	777					1
2.5	2.500	2.517	0.017	10589	0.0 %	10589					-5
5.0	5.000	4.760	-0.240	20194	0.0 %	20194					2
7.5	7.500	7.655	0.155	32590	0.0 %	32590					0
10.0	10.000	9.991	-0.009	42593	0.0 %	42593					

no3
5720/15

LB76128
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0		34	-		0 ISM01.3	PPB	5/20/2015 11:48	Std
0.2 S0.2		777	-		0.2 ISM01.3	PPB	5/20/2015 11:50	Std
2.5 S2.5		10589	-		2.5 ISM01.3	PPB	5/20/2015 11:52	Std
5 S5		20194	-		5 ISM01.3	PPB	5/20/2015 11:54	Std
7.5 S7.5		32590	-		7.5 ISM01.3	PPB	5/20/2015 11:56	Std
10 S10		42593	-		10 ISM01.3	PPB	5/20/2015 11:59	Std
ICV30	ICV30	17545	4.1413	-	ISM01.3	PPB	5/20/2015 12:02	SMPL
ICB30	ICB30	-16	0.0404	-	ISM01.3	PPB	5/20/2015 12:04	SMPL
CCV24	CCV24	20487	4.8284	-	ISM01.3	PPB	5/20/2015 12:06	SMPL
CCB24	CCB24	-16	0.0404	-	ISM01.3	PPB	5/20/2015 12:08	SMPL
PB83484BL	PBW06	-6	0.0427	-	ISM01.3	PPB	5/20/2015 12:10	SMPL
G2212-01	MB0AG4	486	0.1576	-	ISM01.3	PPB	5/20/2015 12:12	SMPL
G2212-02	MB0AG5	-297	-0.0252	-	ISM01.3	PPB	5/20/2015 12:14	SMPL
G2212-03	MB0AG6	-181	0.0019	-	ISM01.3	PPB	5/20/2015 12:16	SMPL
G2212-04	MB0AG7	582	0.1801	-	ISM01.3	PPB	5/20/2015 12:18	SMPL
G2212-05	MB0AG8	-55	0.0313	-	ISM01.3	PPB	5/20/2015 12:21	SMPL
G2212-06	MB0AG9	-159	0.007	-	ISM01.3	PPB	5/20/2015 12:23	SMPL
G2212-07	MB0AH0	725	0.2134	-	ISM01.3	PPB	5/20/2015 12:25	SMPL
G2212-08	MB0AH1	-112	0.018	-	ISM01.3	PPB	5/20/2015 12:27	SMPL
G2212-09	MB0AH1D	-114	0.0175	-	ISM01.3	PPB	5/20/2015 12:29	SMPL
G2212-10	MB0AH1S	4177	1.0196	-	ISM01.3	PPB	5/20/2015 12:31	SMPL
G2212-11	MB0AH3	-173	0.0037	-	ISM01.3	PPB	5/20/2015 12:33	SMPL
G2212-12	MB0AH4	-109	0.0187	-	ISM01.3	PPB	5/20/2015 12:35	SMPL
G2212-13	MB0AH5	-129	0.014	-	ISM01.3	PPB	5/20/2015 12:38	SMPL
PB83486BL	PBW06	-25	0.0383	-	ISM01.3	PPB	5/20/2015 12:40	SMPL
G2302-01	PERF4-INJ-L-20	11320	2.6877	-	ISM01.3	PPB	5/20/2015 12:42	SMPL
G2302-02	PERF4-INJ-L-20-F	113	0.0705	-	ISM01.3	PPB	5/20/2015 12:44	SMPL
G2302-03	PERF4-INJ-L-21	33164	7.7888	-	ISM01.3	PPB	5/20/2015 12:46	SMPL
G2302-04	PERF4-INJ-L-21-F	-209	-0.0047	-	ISM01.3	PPB	5/20/2015 12:48	SMPL
G2302-05	FB051815	-111	0.0182	-	ISM01.3	PPB	5/20/2015 12:50	SMPL
CCV25	CCV25	19966	4.7067	-	ISM01.3	PPB	5/20/2015 12:52	SMPL
CCB25	CCB25	-175	0.0033	-	ISM01.3	PPB	5/20/2015 12:55	SMPL
G2302-06	FB051815-F	-112	0.018	-	ISM01.3	PPB	5/20/2015 12:57	SMPL
G2302-07	RB051815	-79	0.0257	-	ISM01.3	PPB	5/20/2015 12:59	SMPL
G2302-08	RB051815-F	-85	0.0243	-	ISM01.3	PPB	5/20/2015 13:01	SMPL
G2302-09	PERF4-EX-L-5	3917	0.9589	-	ISM01.3	PPB	5/20/2015 13:03	SMPL
G2302-10	G2302-09MS	8002	1.9128	-	ISM01.3	PPB	5/20/2015 13:05	SMPL
G2302-11	G2302-09MSD	3836	0.9399	-	ISM01.3	PPB	5/20/2015 13:07	SMPL
G2302-12	PERF4-EX-L-5-F	-143	0.0107	-	ISM01.3	PPB	5/20/2015 13:09	SMPL
G2302-13	G2302-12MS	4219	1.0294	-	ISM01.3	PPB	5/20/2015 13:12	SMPL
G2302-14	G2302-12MSD	-103	0.0201	-	ISM01.3	PPB	5/20/2015 13:14	SMPL
G2302-15	PERF4-INJ-L-22	11534	2.7376	-	ISM01.3	PPB	5/20/2015 13:16	SMPL
G2302-16	PERF4-INJ-L-22-F	-30	0.0371	-	ISM01.3	PPB	5/20/2015 13:18	SMPL
G2302-17	DUP051815	3165	0.7833	-	ISM01.3	PPB	5/20/2015 13:20	SMPL
G2302-18	DUP051815-F	4	0.0451	-	ISM01.3	PPB	5/20/2015 13:22	SMPL
CCV26	CCV26	19924	4.6969	-	ISM01.3	PPB	5/20/2015 13:24	SMPL

CCB26

CCB26

-190 -0.0002 -

ISM01.3 PPB

5/20/2015 13:26 SMPL
