

MERCURY ANALYSIS LOGBOOK

Date	Case Number	Batch Number	Start Time	BLK	STD1	STD2	STD3	STD4
6/10/15 (1500)	G2376	PB 83935	19:17	145	1063	12193	23796	37042
	G2390	PB 83935						
	G2453	PB 83936						
6/10/15 (1500)	G2555	PB 83954	22:37	119	1063	12237	24595	37361
6/11/15 (1500)	G2521	PB 83947	8:54	101	1188	8946	18601	27211
	G2548							
	G2561							
	G2558	PB 83948						
	G2572							
6/11/15 (1500)	G2556	PB 83949	19:50	46	624	8511	16904	26807
6/12/15 (1500)	G2390	PB 83941	18:29	67	976	11345	22674	31006
	G2453							
	G2513							
	G2484	PB 83962						
	G2551	PB 83963						
6/11/15 (1500)	G2592	PB 83968	14:34	91	743	9146	17546	26678
	G2603							
	G2604							
	G2607							
	G2608							
6/12/15 (1500)	G2590	PB 83974	16:19	232	998	12307	23794	36445
	G2592							
	G2609							
6/15/15 (1500)	G2617	PB 83996	11:52	-17	708	8779	17381	26433
	G2620	PB 83997						
	G2627							
6/16/15 (1500)	G2674	PB 84007	9:30	-2	774	9022	17989	27494
6/16/15 (1500)	G2641	PB 84023	12:48	60	989	11196	21433	30905
6/16/15 (1500)	G2641	PB 84024	14:33	190	1146	1263	25189	38189

MERCURY ANALYSIS LOGBOOK


STD 5	End Time	Correlation Coefficient	Stannous Chloride Prep log #	Comment	Analyst	Supervisor Signature
49496	22:01	0.9997834	MP2397	ICV13-08 CCV13-36-40	AB	AB
					AB	AB
					AB	AB
49682	23:38	0.9999706	MP2797	ICV13-08 CCV13-41-43	AB	AB
35842	10:38	0.9998140	MP2803	ICV13-10 CCV13-44-47	AB	AB
					AB	AB
					AB	AB
					AB	AB
33830	15:01	0.9992682	MP2800	ICV13-11 CCV13-48-50	AB	AB
44010	12:17	0.9992109	MP2803	ICV13-12 CCV13-01-03	AB	AB
					AB	AB
					AB	AB
36988	15:46	0.9994940	MP2803	ICV13-13 CCV13-04-06	AB	AB
					AB	AB
					AB	AB
					AB	AB
47917	17:41	0.9998154	MP2803	ICV13-14 CCV13-07-09	AB	AB
					AB	AB
35941	14:11	0.9998458	MP2803	ICV13-15 CCV13-10-11	AB	AB
					AB	AB
					AB	AB
36854	10:31	0.9999300	MP2809	ICV13-16 CCV13-16-18	AB	AB
42976	13:39	0.9995382	MP2809	ICV13-17 CCV13-09-12	AB	AB
5185	15:58	0.9998539	MP2809	ICV13-18 CCV13-021-023	AB	AB

L378514 CV2

ISM02.2

LB76514
INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.9630e-004

C= -5.3534e-003

Rho= 0.9999539

Accept=Accepted

slope
y=mx+intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.0	0.000	0.032	0.032	190	0.000	190					-
0.2	0.200	0.220	0.020	1146	0.0 %	1146					10
2.5	2.500	2.477	-0.023	12643	0.0 %	12643					-1
5.0	5.000	4.939	-0.061	25189	0.0 %	25189					-1
7.5	7.500	7.490	-0.010	38183	0.0 %	38183					0
10.0	10.000	10.042	0.042	51185	0.0 %	51185					0

N3
6/16/15

LB76514
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	190	-		0 ISM02.2	PPB	6/16/2015 14:33	Std
	0.2 S0.2	1146	-		0.2 ISM02.2	PPB	6/16/2015 14:36	Std
	2.5 S2.5	12643	-		2.5 ISM02.2	PPB	6/16/2015 14:38	Std
	5 S5	25189	-		5 ISM02.2	PPB	6/16/2015 14:40	Std
	7.5 S7.5	38183	-		7.5 ISM02.2	PPB	6/16/2015 14:42	Std
	10 S10	51185	-		10 ISM02.2	PPB	6/16/2015 14:44	Std
ICV	ICV	19834	3.8881 -		ISM02.2	PPB	6/16/2015 14:48	SMPL
ICB	ICB	214	0.0367 -		ISM02.2	PPB	6/16/2015 14:55	SMPL
CCV021	CCV021	25369	4.9747 -		ISM02.2	PPB	6/16/2015 14:57	SMPL
CCB021	CCB021	109	0.016 -		ISM02.2	PPB	6/16/2015 14:59	SMPL
PB84024BL	PBS004	174	0.0288 -		ISM02.2	PPB	6/16/2015 15:02	SMPL
G2641-01	MF9DA0	514	0.0955 -		ISM02.2	PPB	6/16/2015 15:04	SMPL
G2641-02	MF9DA2	1045	0.1998 -		ISM02.2	PPB	6/16/2015 15:06	SMPL
G2641-03	MF9DA3	4	-0.0046 -		ISM02.2	PPB	6/16/2015 15:08	SMPL
G2641-04	MF9DA3D	111	0.0164 -		ISM02.2	PPB	6/16/2015 15:10	SMPL
G2641-05	MF9DA3S	11605	2.2728 -		ISM02.2	PPB	6/16/2015 15:12	SMPL
G2641-12	MF9DA1	-17	-0.0087 -		ISM02.2	PPB	6/16/2015 15:14	SMPL
G2641-13	MF9DA4	103157	20.2448 -		ISM02.2	PPB	6/16/2015 15:16	SMPL
G2641-14	MF9DA5	-97	-0.0244 -		ISM02.2	PPB	6/16/2015 15:18	SMPL
G2641-15	MF9DA6	546	0.1018 -		ISM02.2	PPB	6/16/2015 15:21	SMPL
G2641-16	MF9DA7	201	0.0341 -		ISM02.2	PPB	6/16/2015 15:23	SMPL
G2641-17	MF9DA8	42613	8.3597 -		ISM02.2	PPB	6/16/2015 15:25	SMPL
G2641-18	MF9DA9	143	0.0227 -		ISM02.2	PPB	6/16/2015 15:27	SMPL
G2641-19	MF9DB0	1031	0.197 -		ISM02.2	PPB	6/16/2015 15:29	SMPL
G2641-20	MF9DB1	134	0.021 -		ISM02.2	PPB	6/16/2015 15:31	SMPL
G2641-21	MF9DB2	54121	10.6188 -		ISM02.2	PPB	6/16/2015 15:33	SMPL
G2641-22	MF9DB3	731	0.1381 -		ISM02.2	PPB	6/16/2015 15:35	SMPL
G2641-23	MF9DB4	44876	8.804 -		ISM02.2	PPB	6/16/2015 15:38	SMPL
G2641-24	MF9DB5	47534	9.3258 -		ISM02.2	PPB	6/16/2015 15:40	SMPL
CCV022	CCV022	24486	4.8013 -		ISM02.2	PPB	6/16/2015 15:42	SMPL
CCB022	CCB022	-43	-0.0138 -		ISM02.2	PPB	6/16/2015 15:44	SMPL
G2641-13DLX5	MF9DA4	19455	3.8137 -		ISM02.2	PPB	6/16/2015 15:50	SMPL
G2641-21DLX5	MF9DB2	9500	1.8595 -		ISM02.2	PPB	6/16/2015 15:52	SMPL
CCV023	CCV023	24593	4.8223 -		ISM02.2	PPB	6/16/2015 15:54	SMPL
CCB023	CCB023	90	0.0123 -		ISM02.2	PPB	6/16/2015 15:56	SMPL