

MERCURY ANALYSIS LOGBOOK[illegible]

QA Control # A3041058

Page 58

MERCURY ANALYSIS LOGBOOK

[illegible]

QA Control # A3041058

Page 59

LB 76520 CVI

ISM01.3


LB76526

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39615



A= 0.0000e+000

B= 2.5690e-004

C= -5.3333e-002

Rho= 0.9997425

Accept=Accepted

Slope
 $y = 1.116 \times 10^{-4} x$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	90 D
0.0	0.000	0.007	0.007	236	0.000	236					-
0.2	0.200	0.245	0.045	1163	0.0 %	1163					23
2.5	2.500	2.530	0.030	10057	0.0 %	10057					1
5.0	5.000	4.926	-0.074	19384	0.0 %	19384					-1
7.5	7.500	7.366	-0.134	28879	0.0 %	28879					-2
10.0	10.000	10.125	0.125	39619	0.0 %	39619					1

NB
6/16/15

LB76520
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0	236	-		0 ISM01.3	PPB	6/16/2015 17:35	Std
	0.2 S0.2	1163	-		0.2 ISM01.3	PPB	6/16/2015 17:37	Std
	2.5 S2.5	10057	-		2.5 ISM01.3	PPB	6/16/2015 17:39	Std
	5 S5	19384	-		5 ISM01.3	PPB	6/16/2015 17:42	Std
	7.5 S7.5	28879	-		7.5 ISM01.3	PPB	6/16/2015 17:44	Std
	10 S10	39619	-		10 ISM01.3	PPB	6/16/2015 17:49	Std
ICV17	ICV17	17111	4.3425 -		ISM01.3	PPB	6/16/2015 17:52	SMPL
ICB17	ICB17	161	-0.012 -		ISM01.3	PPB	6/16/2015 17:57	SMPL
CCV24	CCV24	19559	4.9714 -		ISM01.3	PPB	6/16/2015 17:59	SMPL
CCB24	CCB24	143	-0.0166 -		ISM01.3	PPB	6/16/2015 18:01	SMPL
PB84044BL	PBW06	176	-0.0081 -		ISM01.3	PPB	6/16/2015 18:03	SMPL
G2643-02	MA4G03	157	-0.013 -		ISM01.3	PPB	6/16/2015 18:05	SMPL
G2643-03	MA4G04	192	-0.004 -		ISM01.3	PPB	6/16/2015 18:07	SMPL
G2643-04	MA4G04D	150	-0.0148 -		ISM01.3	PPB	6/16/2015 18:09	SMPL
G2643-05	MA4G04S	4470	1.095 -		ISM01.3	PPB	6/16/2015 18:12	SMPL
G2643-06	MA4G07	589	0.098 -		ISM01.3	PPB	6/16/2015 18:14	SMPL
G2643-07	MA4G08	80	-0.0328 -		ISM01.3	PPB	6/16/2015 18:16	SMPL
G2643-08	MA4G10	101	-0.0274 -		ISM01.3	PPB	6/16/2015 18:18	SMPL
G2643-11	MA4G00	16418	4.1645 -		ISM01.3	PPB	6/16/2015 18:20	SMPL
G2643-12	MA4G12	94	-0.0292 -		ISM01.3	PPB	6/16/2015 18:22	SMPL
G2643-13	MA4G13	97	-0.0284 -		ISM01.3	PPB	6/16/2015 18:24	SMPL
G2643-14	MA4G14	120	-0.0225 -		ISM01.3	PPB	6/16/2015 18:26	SMPL
G2643-15	MA4G19	120	-0.0225 -		ISM01.3	PPB	6/16/2015 18:29	SMPL
G2643-16	MA4G21	174	-0.0086 -		ISM01.3	PPB	6/16/2015 18:31	SMPL
G2643-17	MA4G22	90	-0.0302 -		ISM01.3	PPB	6/16/2015 18:33	SMPL
PB84045BL	PBW06	122	-0.022 -		ISM01.3	PPB	6/16/2015 18:35	SMPL
G2655-01	MA4FZ1	132	-0.0194 -		ISM01.3	PPB	6/16/2015 18:37	SMPL
G2655-02	MA4G05	482	0.0705 -		ISM01.3	PPB	6/16/2015 18:39	SMPL
G2655-03	MA4G06	277	0.0178 -		ISM01.3	PPB	6/16/2015 18:41	SMPL
G2655-04	MA4G09	-38	-0.0631 -		ISM01.3	PPB	6/16/2015 18:43	SMPL
CCV25	CCV25	18079	4.5912 -		ISM01.3	PPB	6/16/2015 18:46	SMPL
CCB25	CCB25	15	-0.0495 -		ISM01.3	PPB	6/16/2015 18:48	SMPL
G2655-05	MA4G15	74	-0.0343 -		ISM01.3	PPB	6/16/2015 18:50	SMPL
G2655-06	MA4G15D	86	-0.0312 -		ISM01.3	PPB	6/16/2015 18:52	SMPL
G2655-07	MA4G15S	4124	1.0061 -		ISM01.3	PPB	6/16/2015 18:54	SMPL
G2655-08	MA4G16	46	-0.0415 -		ISM01.3	PPB	6/16/2015 18:56	SMPL
CCV26	CCV26	18535	4.7083 -		ISM01.3	PPB	6/16/2015 19:05	SMPL
CCB26	CCB26	-38	-0.0631 -		ISM01.3	PPB	6/16/2015 19:07	SMPL