

MERCURY ANALYSIS LOGBOOK

Date	Case Number	Batch Number	Start Time	BLK	STD1	STD2	STD3	STD4
6/25/15 (7420)	Q2720	PB 84191	10:54	-36	678	9100	17701	26751
↓	Q2748	↓	↓	↓	↓	↓	↓	↓
6/25/15 (7421)	Q2747	PB 84194	12:28	52	922	12434	24434	35918
↓	Q2749	↓	↓	↓	↓	↓	↓	↓
↓	Q2750	↓	↓	↓	↓	↓	↓	↓
6/25/15 (15M0)	Q2754	PB 84192	14:26	125	1092	12225	24486	36844
↓	Q2755	PB 84193	↓	↓	↓	↓	↓	↓
↓	Q2756	PB 84202	↓	↓	↓	↓	↓	↓

MERCURY ANALYSIS LOGBOOK

[illegible]

LB 76647 CV1


ISM02.2

LB76647
INSTRUMENT ID: CV1

Linear

 μ Abs.:

49781



A= 0.0000e+000

B= 2.0213e-004

C= 2.3374e-003

Rho= 0.9999351

Accept=Accepted

slope
y=intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.028	0.028	125	0.000	125					—
0.2	0.200	0.223	0.023	1092	0.0 %	1092					12
2.5	2.500	2.483	-0.017	12275	0.0 %	12275					—1
5.0	5.000	4.952	-0.048	24486	0.0 %	24486					—1
7.5	7.500	7.450	-0.050	36844	0.0 %	36844					—1
10.0	10.000	10.065	0.065	49781	0.0 %	49781					1

LB
6/25/11

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	125	-		0 ISM02.2	PPB	6/25/2015 14:26	Std
	0.2 S0.2	1092	-		0.2 ISM02.2	PPB	6/25/2015 14:28	Std
	2.5 S2.5	12275	-		2.5 ISM02.2	PPB	6/25/2015 14:30	Std
	5 S5	24486	-		5 ISM02.2	PPB	6/25/2015 14:32	Std
	7.5 S7.5	36844	-		7.5 ISM02.2	PPB	6/25/2015 14:34	Std
	10 S10	49781	-		10 ISM02.2	PPB	6/25/2015 14:36	Std
ICV	ICV	18977	3.8382	-	ISM02.2	PPB	6/25/2015 14:39	SMPL
ICB	ICB	43	0.011	-	ISM02.2	PPB	6/25/2015 14:42	SMPL
CCV023	CCV023	24082	4.87	-	ISM02.2	PPB	6/25/2015 14:44	SMPL
CCB023	CCB023	34	0.0092	-	ISM02.2	PPB	6/25/2015 14:46	SMPL
PB84192BL	PBS004	79	0.0183	-	ISM02.2	PPB	6/25/2015 14:48	SMPL
G2754-01	MC00S3	22825	4.6159	-	ISM02.2	PPB	6/25/2015 14:50	SMPL
G2754-02	MC00S4	55362	11.1926	-	ISM02.2	PPB	6/25/2015 14:52	SMPL
G2754-03	MC00T7	3956	0.802	-	ISM02.2	PPB	6/25/2015 14:54	SMPL
G2754-04	MC00W0	1561	0.3179	-	ISM02.2	PPB	6/25/2015 14:56	SMPL
G2754-05	MC00WOD	1467	0.2989	-	ISM02.2	PPB	6/25/2015 14:58	SMPL
G2754-06	MC00W0S	13210	2.6725	-	ISM02.2	PPB	6/25/2015 15:01	SMPL
G2754-07	MC00W1	7491	1.5165	-	ISM02.2	PPB	6/25/2015 15:03	SMPL
G2754-08	MC00W2	16811	3.4003	-	ISM02.2	PPB	6/25/2015 15:05	SMPL
G2754-09	MC00W4	2155	0.4379	-	ISM02.2	PPB	6/25/2015 15:07	SMPL
G2754-10	MC00Z5	7045	1.4263	-	ISM02.2	PPB	6/25/2015 15:09	SMPL
G2754-11	MC00Z6	6325	1.2808	-	ISM02.2	PPB	6/25/2015 15:11	SMPL
G2754-12	MC00Z7	5923	1.1996	-	ISM02.2	PPB	6/25/2015 15:13	SMPL
G2754-13	MC00Z8	5957	1.2064	-	ISM02.2	PPB	6/25/2015 15:15	SMPL
G2754-14	MC00Z9	6075	1.2303	-	ISM02.2	PPB	6/25/2015 15:18	SMPL
G2754-15	MC01A0	7678	1.5543	-	ISM02.2	PPB	6/25/2015 15:20	SMPL
G2754-16	MC01A1	6987	1.4146	-	ISM02.2	PPB	6/25/2015 15:22	SMPL
G2754-17	MC01A2	6172	1.2499	-	ISM02.2	PPB	6/25/2015 15:24	SMPL
G2754-18	MC01A3	8174	1.6545	-	ISM02.2	PPB	6/25/2015 15:26	SMPL
G2754-19	MC01A4	5476	1.1092	-	ISM02.2	PPB	6/25/2015 15:28	SMPL
CCV024	CCV024	23348	4.7217	-	ISM02.2	PPB	6/25/2015 15:30	SMPL
CCB024	CCB024	-164	-0.0308	-	ISM02.2	PPB	6/25/2015 15:32	SMPL
G2754-20	MC01A5	7507	1.5197	-	ISM02.2	PPB	6/25/2015 15:35	SMPL
G2754-21	MC01A7	4471	0.9061	-	ISM02.2	PPB	6/25/2015 15:37	SMPL
G2754-22	MC01B2	30080	6.0824	-	ISM02.2	PPB	6/25/2015 15:39	SMPL
PB84193BL	PBS004	-262	-0.0506	-	ISM02.2	PPB	6/25/2015 15:41	SMPL
G2755-01	MC00R1	6668	1.3501	-	ISM02.2	PPB	6/25/2015 15:43	SMPL
G2755-02	MC00R2	753	0.1545	-	ISM02.2	PPB	6/25/2015 15:45	SMPL
G2755-03	MC00R2D	808	0.1657	-	ISM02.2	PPB	6/25/2015 15:50	SMPL
G2755-04	MC00R2S	12260	2.4804	-	ISM02.2	PPB	6/25/2015 15:53	SMPL
G2755-05	MC00R5	11804	2.3883	-	ISM02.2	PPB	6/25/2015 15:55	SMPL
G2755-06	MC00R6	3463	0.7023	-	ISM02.2	PPB	6/25/2015 15:57	SMPL
G2755-07	MC00R7	3233	0.6558	-	ISM02.2	PPB	6/25/2015 15:59	SMPL
G2755-08	MC00R8	1041	0.2128	-	ISM02.2	PPB	6/25/2015 16:01	SMPL
G2755-09	MC00T8	7049	1.4272	-	ISM02.2	PPB	6/25/2015 16:03	SMPL
G2755-10	MC00T9	3685	0.7472	-	ISM02.2	PPB	6/25/2015 16:05	SMPL

INSTRUMENT ID : CV1

G2755-11	MC00W3	1462	0.2979 -	ISM02.2	PPB	6/25/2015 16:07 SMPL
G2755-12	MC00Z1	9677	1.9583 -	ISM02.2	PPB	6/25/2015 16:10 SMPL
G2755-13	MC00Z2	6991	1.4154 -	ISM02.2	PPB	6/25/2015 16:12 SMPL
G2755-14	MC00Z3	5718	1.1581 -	ISM02.2	PPB	6/25/2015 16:14 SMPL
G2755-15	MC00Z4	8075	1.6345 -	ISM02.2	PPB	6/25/2015 16:16 SMPL
G2755-16	MC00Y1	4478	0.9075 -	ISM02.2	PPB	6/25/2015 16:18 SMPL
CCV025	CCV025	23109	4.6734 -	ISM02.2	PPB	6/25/2015 16:20 SMPL
CCB025	CCB025	-135	-0.025 -	ISM02.2	PPB	6/25/2015 16:22 SMPL
G2755-17	MC00Y2	2700	0.5481 -	ISM02.2	PPB	6/25/2015 16:24 SMPL
G2755-18	MC00Y3	3882	0.787 -	ISM02.2	PPB	6/25/2015 16:27 SMPL
G2755-19	MC00Y4	4674	0.9471 -	ISM02.2	PPB	6/25/2015 16:29 SMPL
G2755-20	MC00Y5	4991	1.0112 -	ISM02.2	PPB	6/25/2015 16:31 SMPL
G2755-21	MC00Y6	2866	0.5816 -	ISM02.2	PPB	6/25/2015 16:33 SMPL
G2755-22	MC00Y7	4814	0.9754 -	ISM02.2	PPB	6/25/2015 16:35 SMPL
PB84202BL	PBS004	-66	-0.011 -	ISM02.2	PPB	6/25/2015 16:37 SMPL
G2756-01	MC00Y8	7232	1.4641 -	ISM02.2	PPB	6/25/2015 16:39 SMPL
G2756-02	MC00Y9	9270	1.8761 -	ISM02.2	PPB	6/25/2015 16:41 SMPL
G2756-03	MC00Z0	14633	2.9601 -	ISM02.2	PPB	6/25/2015 16:44 SMPL
G2756-04	MC01A6	7894	1.5979 -	ISM02.2	PPB	6/25/2015 16:46 SMPL
G2756-05	MC00Q9	24969	5.0493 -	ISM02.2	PPB	6/25/2015 16:48 SMPL
G2756-06	MC00R0	31031	6.2746 -	ISM02.2	PPB	6/25/2015 16:50 SMPL
G2756-07	MC00R3	5656	1.1456 -	ISM02.2	PPB	6/25/2015 16:52 SMPL
G2756-08	MC00R4	5845	1.1838 -	ISM02.2	PPB	6/25/2015 16:54 SMPL
G2756-09	MC00R9	5328	1.0793 -	ISM02.2	PPB	6/25/2015 16:56 SMPL
G2756-10	MC00S0	1110	0.2267 -	ISM02.2	PPB	6/25/2015 16:58 SMPL
G2756-11	MC00S1	2086	0.424 -	ISM02.2	PPB	6/25/2015 17:01 SMPL
G2756-12	MC00S2	3079	0.6247 -	ISM02.2	PPB	6/25/2015 17:03 SMPL
G2756-13	MC00S5	13320	2.6947 -	ISM02.2	PPB	6/25/2015 17:05 SMPL
CCV026	CCV026	23282	4.7083 -	ISM02.2	PPB	6/25/2015 17:07 SMPL
CCB026	CCB026	-96	-0.0171 -	ISM02.2	PPB	6/25/2015 17:09 SMPL
G2756-14	MC00S6	5313	1.0763 -	ISM02.2	PPB	6/25/2015 17:11 SMPL
G2756-15	MC00S6D	4982	1.0093 -	ISM02.2	PPB	6/25/2015 17:13 SMPL
G2756-16	MC00S6S	16542	3.346 -	ISM02.2	PPB	6/25/2015 17:15 SMPL
G2756-17	MC00S7	17962	3.633 -	ISM02.2	PPB	6/25/2015 17:18 SMPL
G2756-18	MC00S8	16982	3.4349 -	ISM02.2	PPB	6/25/2015 17:20 SMPL
G2756-19	MC00S9	4250	0.8614 -	ISM02.2	PPB	6/25/2015 17:22 SMPL
G2756-20	MC00T0	2157	0.4383 -	ISM02.2	PPB	6/25/2015 17:24 SMPL
G2756-21	MC00T1	13488	2.7287 -	ISM02.2	PPB	6/25/2015 17:26 SMPL
G2756-22	MC00T2	3438	0.6973 -	ISM02.2	PPB	6/25/2015 17:28 SMPL
G2754-02DLX5	MC00S4	9646	1.9521 -	ISM02.2	PPB	6/25/2015 17:33 SMPL
CCV027	CCV027	23208	4.6934 -	ISM02.2	PPB	6/25/2015 17:35 SMPL
CCB027	CCB027	-38	-0.0053 -	ISM02.2	PPB	6/25/2015 17:37 SMPL