


ISM02.2

LB 76892
INSTRUMENT ID: CV1

Linear

 μ Abs.:

39850



A= 0.0000e+000

B= 2.4788e-004

C= 3.5029e-002

Rho= 0.9993848

Accept=Accepted

slope
 $y = mx + b$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.055	0.055	79	0.000	79					-
0.2	0.200	0.243	0.043	838	0.0 %	838					22
2.5	2.500	2.410	-0.090	9583	0.0 %	9583					-4
5.0	5.000	4.844	-0.156	19399	0.0 %	19399					-3
7.5	7.500	7.736	0.236	31067	0.0 %	31067					3
10.0	10.000	9.913	-0.087	39850	0.0 %	39850					-1

LB76892
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	79	-		0 ISM02.2	PPB	7/14/2015 16:04	Std
	0.2 S0.2	838	-		0.2 ISM02.2	PPB	7/14/2015 16:06	Std
	2.5 S2.5	9583	-		2.5 ISM02.2	PPB	7/14/2015 16:08	Std
	5 S5	19399	-		5 ISM02.2	PPB	7/14/2015 16:12	Std
	7.5 S7.5	31067	-		7.5 ISM02.2	PPB	7/14/2015 16:16	Std
	10 S10	39850	-		10 ISM02.2	PPB	7/14/2015 16:18	Std
ICV	ICV	15643	3.9125 -		ISM02.2	PPB	7/14/2015 16:21	SMPL
ICB	ICB	45	0.0462 -		ISM02.2	PPB	7/14/2015 16:23	SMPL
CCV046	CCV046	19251	4.8069 -		ISM02.2	PPB	7/14/2015 16:25	SMPL
CCB046	CCB046	39	0.0447 -		ISM02.2	PPB	7/14/2015 16:27	SMPL
PB84457BL	PBW003	88	0.0568 -		ISM02.2	PPB	7/14/2015 16:29	SMPL
G2952-01	MJG5Y0	39	0.0447 -		ISM02.2	PPB	7/14/2015 16:31	SMPL
G2952-02	MJG5Y0D	-43	0.0244 -		ISM02.2	PPB	7/14/2015 16:33	SMPL
G2952-03	MJG5Y0S	3741	0.9623 -		ISM02.2	PPB	7/14/2015 16:35	SMPL
G2952-04	MJG5Y1	-181	-0.0098 -		ISM02.2	PPB	7/14/2015 16:38	SMPL
G2952-05	MJG5Y2	-33	0.0268 -		ISM02.2	PPB	7/14/2015 16:40	SMPL
G2952-06	MJG5Y3	-45	0.0239 -		ISM02.2	PPB	7/14/2015 16:42	SMPL
G2952-07	MJG5Y4	-21	0.0298 -		ISM02.2	PPB	7/14/2015 16:44	SMPL
G2952-08	MJG5Y5	0	0.035 -		ISM02.2	PPB	7/14/2015 16:46	SMPL
CCV047	CCV047	18659	4.6601 -		ISM02.2	PPB	7/14/2015 16:48	SMPL
CCB047	CCB047	-68	0.0182 -		ISM02.2	PPB	7/14/2015 16:50	SMPL