

RS 04-18-19


~~L B102167~~ L B102175

ISM02.4

INSTRUMENT ID : CV1

μ Abs.:

4518E



Linear

A= 0.0000e+000

B= 2.2008e-004

C= -3.5619e-002

Rho= 0.9998976

Accept=Accepted

*slope
y-intercept*

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.021	-0.021	65	0.000	65					-
0.05	0.050	0.053	0.003	403	0.0 %	403					6
0.20	0.200	0.163	-0.037	904	0.0 %	904					-19
2.50	2.500	2.509	0.009	11561	0.0 %	11561					0
5.00	5.000	5.061	0.061	23158	0.0 %	23158					1
7.50	7.500	7.576	0.076	34587	0.0 %	34587					-1
10.0	10.000	9.909	-0.091	45185	0.0 %	45185					-1

RS 04-17-19

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	65	-		0 ISM02.4	PPB	4/17/2019 17:38	Std
	0.05 S0.05	403	-	0.05	ISM02.4	PPB	4/17/2019 17:43	Std
	0.2 S0.2	904	-	0.2	ISM02.4	PPB	4/17/2019 17:46	Std
	2.5 S2.5	11561	-	2.5	ISM02.4	PPB	4/17/2019 17:48	Std
	5 S5.0	23158	-	5	ISM02.4	PPB	4/17/2019 17:50	Std
	7.5 S7.5	34587	-	7.5	ISM02.4	PPB	4/17/2019 17:52	Std
	10 S.10.0	45185	-	10	ISM02.4	PPB	4/17/2019 17:55	Std
ICV	ICV	17205	3.7509	-	ISM02.4	PPB	4/17/2019 17:58	SMPL
ICB	ICB	0	-0.0356	-	ISM02.4	PPB	4/17/2019 18:00	SMPL
CCV028	CCV028	23143	5.0577	-	ISM02.4	PPB	4/17/2019 18:02	SMPL
CCB028	CCB028	0	-0.0356	-	ISM02.4	PPB	4/17/2019 18:04	SMPL
PB118978BL	PBW978	21	-0.031	-	ISM02.4	PPB	4/17/2019 18:07	SMPL
K2350-01	METC19	21	-0.031	-	ISM02.4	PPB	4/17/2019 18:09	SMPL
K2350-02	METC19D	15	-0.0323	-	ISM02.4	PPB	4/17/2019 18:11	SMPL
K2350-03	METC19S	4889	1.0404	-	ISM02.4	PPB	4/17/2019 18:13	SMPL
CCV029	CCV029	23063	5.0401	-	ISM02.4	PPB	4/17/2019 18:52	SMPL
CCB029	CCB029	27	-0.0297	-	ISM02.4	PPB	4/17/2019 18:54	SMPL