


LB103851

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.006	0.006	50	0.000	50				
0.2	0.200	0.256	0.056	1067	0.0 %	1067				
2.5	2.500	2.402	-0.098	9789	0.0 %	9789				
5.0	5.000	4.985	-0.015	20291	0.0 %	20291				
7.5	7.500	7.590	0.090	30881	0.0 %	30881				
10.0	10.000	9.962	-0.038	40524	0.0 %	40524				

LB103851 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 Jul 2019 11:10:30

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	50	-	0.00007470A	PPB		09 Jul 2019 11:12:03	S	Std
0.2 - 1	50.2	1067	-	0.20007470A	PPB		09 Jul 2019 11:19:35	S	Std
2.5 - 1	52.5	9789	-	2.50007470A	PPB		09 Jul 2019 11:21:43	S	Std
5.0 - 1	55.0	20291	-	5.00007470A	PPB		09 Jul 2019 11:23:52	S	Std
7.5 - 1	57.5	30881	-	7.50007470A	PPB		09 Jul 2019 11:26:00	S	Std
10.0 - 1	510.0	40524	-	10.00007470A	PPB		09 Jul 2019 11:28:06	S	Std
ICV24 - 1	ICV24	16988	4.1724	-7470A	PPB		09 Jul 2019 11:32:57	U	SMPL
ICB24 - 1	ICB24	125	0.0245	-7470A	PPB		09 Jul 2019 11:44:58	U	SMPL
CCV99 - 1	CCV99	20299	4.9868	-7470A	PPB		09 Jul 2019 11:47:04	U	SMPL
CCB99 - 1	CCB99	140	0.0282	-7470A	PPB		09 Jul 2019 11:53:33	U	SMPL
CRA - 1	CRA	1052	0.2526	-7470A	PPB		09 Jul 2019 11:55:44	U	SMPL
HighStd - 1	HighStd	40558	9.9699	-7470A	PPB		09 Jul 2019 11:57:57	U	SMPL
ChkStd - 1	ChkStd	28053	6.8940	-7470A	PPB		09 Jul 2019 12:00:09	U	SMPL
PB121237BL - 1	PBW	-51	-0.0187	-7470A	PPB		09 Jul 2019 12:02:19	U	SMPL
PB121237BS - 1	LCSW	15352	3.7700	-7470A	PPB		09 Jul 2019 12:04:27	U	SMPL
K3666-01 - 1	MWHR3-14-070219	117	0.0226	-7470A	PPB		09 Jul 2019 12:06:35	U	SMPL
K3666-01DUP - 1	MWHR3-14-070219DUP	158	0.0327	-7470A	PPB		09 Jul 2019 12:08:43	U	SMPL
K3666-01MS - 1	MWHR3-14-070219MS	15441	3.7919	-7470A	PPB		09 Jul 2019 12:10:51	U	SMPL
K3666-01MSD - 1	MWHR3-14-070219MSD	15344	3.7680	-7470A	PPB		09 Jul 2019 12:13:00	U	SMPL
K3666-02 - 1	MWHR3-8-070219	631	0.1490	-7470A	PPB		09 Jul 2019 12:15:07	U	SMPL
GCV400 - 1 CCV01	GCV400 CCV01	19650	4.8271	-7470A	PPB		09 Jul 2019 12:17:14	U	SMPL
CCB400 - 1 CCB01	CCB400 CCB01	77	0.0127	-7470A	PPB		09 Jul 2019 12:19:20	U	SMPL
K3666-03 - 1	MWHR2-8-070219	776	0.1847	-7470A	PPB		09 Jul 2019 12:21:27	U	SMPL
K3666-04 - 1	MWHR2-9-070219	742	0.1763	-7470A	PPB		09 Jul 2019 12:23:35	U	SMPL
K3704-01 - 1	EFFL-B2-9	731	0.1736	-7470A	PPB		09 Jul 2019 12:25:42	U	SMPL
K3666-01LX5 - 1		14	-0.0028	-7470A	PPB		09 Jul 2019 12:33:22	U	SMPL
K3666-01A - 1		15299	3.7569	-7470A	PPB		09 Jul 2019 12:35:29	U	SMPL
GCV401 - 1 CCV02	GCV401 CCV02	19730	4.8468	-7470A	PPB		09 Jul 2019 12:37:37	U	SMPL
CCB401 - 1 CCB02	CCB401 CCB02	-37	-0.0153	-7470A	PPB		09 Jul 2019 12:39:45	U	SMPL

RS 07-09-19