

MERCURY ANALYSIS LOGBOOK

[illegible]

QA Control # A3040930

Page 50


MERCURY ANALYSIS LOGBOOK

[illegible]

QA Control # A3040930

Page: 51

LB 59492 A-B



Instrument ID: CR2 LB59492 A-P.

Line	Conc.	Units	SD/RSD	1	2	3	4	5

*** Standard: 1 Rep: 1			50	Seq: 1		13:05:33 24 Feb 12	HG	
Hg	.000	ppb	4606					
*** Standard: 2 Rep: 1			50.2	Seq: 2		13:07:36 24 Feb 12	HG	
Hg	.200	ppb	16745					
*** Standard: 3 Rep: 1			52.5	Seq: 3		13:09:48 24 Feb 12	HG	
Hg	2.50	ppb	193532					
*** Standard: 4 Rep: 1			55.0	Seq: 4		13:11:49 24 Feb 12	HG	
Hg	5.00	ppb	349946					
*** Standard: 5 Rep: 1			57.5	Seq: 5		13:13:58 24 Feb 12	HG	
Hg	7.50	ppb	556793					
*** Standard: 6 Rep: 1			510.0	Seq: 6		13:16:15 24 Feb 12	HG	
Hg	10.0	ppb	712535					

02/24/12 A-P.

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: ICV								
			ICV	Seq: 8	13:22:23 24 Feb 12	HG		
Hg	3.90	ppb	.000	3.90				
*** Sample ID: ICB								
			ICB	Seq: 9	13:24:26 24 Feb 12	HG		
Hg	-.119	ppb	.000	-.119				
*** Check Standard: 2 Ck2								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		103.	5.13	5.00	ppb	.000		
*** Check Standard: 1 Ck1								
				Seq: 11	13:28:52 24 Feb 12	HG		
Line	Flag	Found	Range(+/-)	Units	SD/RSD			
Hg		.118	.200	ppb	.000			
*** Sample ID: CRA								
			CRA	Seq: 12	13:31:00 24 Feb 12	HG		
Hg	.526	ppb	.000	.526				
*** Sample ID: HIGH STD								
			HIGH STD	Seq: 13	13:33:02 24 Feb 12	HG		
Hg	9.60	ppb	.000	9.60				
*** Sample ID: CHK STD								
			CHK STD	Seq: 14	13:35:43 24 Feb 12	HG		
Hg	6.86	ppb	.000	6.86				
*** Sample ID: PB61343BL <i>2222</i> <i>A-P 02/24/12</i>								
			PBS	Seq: 15	13:37:41 24 Feb 12	HG		
Hg	.285	ppb	.000	.285				
*** Sample ID: PB61343BS								
			LCSS	Seq: 16	13:39:45 24 Feb 12	HG		
Hg	3.83	ppb	.000	3.83				
*** Sample ID: D1576-01								
				Seq: 17	13:41:48 24 Feb 12	HG		
Hg	1.99	ppb	.000	1.99				
*** Sample ID: D1576-02								
				Seq: 18	13:44:30 24 Feb 12	HG		
Hg	4.50	ppb	.000	4.50				
*** Sample ID: D1576-03								
				Seq: 19	13:46:31 24 Feb 12	HG		
Hg	1.60	ppb	.000	1.60				

Line	Conc.	Units	SD/RSD		1	2	3	4	5

***	Check Standard: 2		Ck2		Seq:	20	13:48:32	24 Feb 12	HG
Line	Flag	%Rcv.	Found	True	Units		SD/RSD		
Hg		98.7	4.94	5.00	ppb		.000		
***	Check Standard: 1		Ck1		Seq:	21	13:50:36	24 Feb 12	HG
Line	Flag	Found	Range(+/-)	Units		SD/RSD			
Hg		-.122	.200	ppb		.000			
***	Sample ID: D1576-04				Seq:	22	13:52:41	24 Feb 12	HG
Hg	2.50	ppb	.000		2.50				
***	Sample ID: D1576-05				Seq:	23	13:54:51	24 Feb 12	HG
Hg	2.32	ppb	.000		2.32				
***	Sample ID: D1576-06				Seq:	24	13:56:55	24 Feb 12	HG
Hg	2.62	ppb	.000		2.62				
***	Sample ID: D1576-07				Seq:	25	13:58:54	24 Feb 12	HG
Hg	.930	ppb	.000		.930				
***	Sample ID: D1577-01				Seq:	26	14:00:58	24 Feb 12	HG
Hg	-.088	ppb	.000		-.088				
***	Sample ID: D1577-01D				Seq:	27	14:03:18	24 Feb 12	HG
Hg	.138	ppb	.000		.138				
***	Sample ID: D1577-01L				Seq:	28	14:05:27	24 Feb 12	HG
Hg	.095	ppb	.000		.095				
***	Sample ID: D1577-02				Seq:	29	14:07:47	24 Feb 12	HG
Hg	4.07	ppb	.000		4.07				
***	Check Standard: 2		Ck2		Seq:	30	14:10:08	24 Feb 12	HG
Line	Flag	%Rcv.	Found	True	Units		SD/RSD		
Hg		96.5	4.82	5.00	ppb		.000		
***	Check Standard: 1		Ck1		Seq:	31	14:12:11	24 Feb 12	HG
Line	Flag	Found	Range(+/-)	Units		SD/RSD			
Hg		-.123	.200	ppb		.000			

Protocol: chemtech

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
------	-------	-------	--------	---	---	---	---	---

*** Sample ID: D1577-03				Seq: 32		14:14:31	24 Feb 12	HG
Hg	4.16	ppb	.000	4.16				
=====								
*** Sample ID: D1577-01A				Seq: 33		14:16:42	24 Feb 12	HG
Hg	4.07	ppb	.000	4.07				
=====								
*** Sample ID: D1577-04				Seq: 34		14:18:42	24 Feb 12	HG
Hg	.208	ppb	.000	.208				
=====								
*** Sample ID: D1578-01				Seq: 35		14:20:42	24 Feb 12	HG
Hg	3.83	ppb	.000	3.83				
=====								
*** Sample ID: D1578-02				Seq: 36		14:22:44	24 Feb 12	HG
Hg	2.29	ppb	.000	2.29				
=====								
*** Sample ID: D1596-01				Seq: 37		14:25:13	24 Feb 12	HG
Hg	-.013	ppb	.000	-.013				
=====								
*** Sample ID: D1598-01				Seq: 38		14:27:14	24 Feb 12	HG
Hg	.057	ppb	.000	.057				
=====								
*** Sample ID: PB61343BL				Seq: 39		14:29:34	24 Feb 12	HG
Hg	-.007	ppb	.000	-.007				
=====								
*** Check Standard: 2 Ck2				Seq: 40		14:31:58	24 Feb 12	HG
Line Flag %Rcv. Found True Units SD/RSD								
Hg		101.	5.05	5.00	ppb	.000		
=====								
*** Check Standard: 1 Ck1				Seq: 44		14:42:48	24 Feb 12	HG
Line Flag Found Range(+/-) Units SD/RSD								
Hg		.197	.200		ppb	.000		
=====								