


MERCURY ANALYSIS LOGBOOK

Date	Case Number	Batch Number	Start Time	BLK	STD1	STD2	STD3	STD4
03/06/12	D1683	PB61525	12:13	12104	24345	118542	361415	52966
	D1684	↓						
	D1686	PB61522						
	D1687	PB61525						
	D1687	PB61524						
	D1688	PB61526						
	D1687	PB61525						
	D1707	PB61524						
03/12/12	D1753	PB61616	9:29	8541	19216	137570	281840	37806
	D1754	↓						
	D1755							

MERCURY ANALYSIS LOGBOOK

[illegible]

LB 59684 A-P.



Instrument ID: C12 LB59684 A-P.

Protocol: chemtech

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1								
			50					
Hg	.000	ppb	8541	Seq: 1		09:29:13	12 Mar 12	HG
*** Standard: 2 Rep: 1								
			50.2					
Hg	.200	ppb	19216	Seq: 2		09:31:14	12 Mar 12	HG
*** Standard: 3 Rep: 1								
			52.5					
Hg	2.50	ppb	137810	Seq: 3		09:33:25	12 Mar 12	HG
*** Standard: 4 Rep: 1								
			55.0					
Hg	5.00	ppb	281840	Seq: 4		09:35:53	12 Mar 12	HG
*** Standard: 5 Rep: 1								
			57.5					
Hg	7.50	ppb	377806	Seq: 5		09:38:09	12 Mar 12	HG
*** Standard: 6 Rep: 1								
			510.0					
Hg	10.0	ppb	501082	Seq: 6		09:40:26	12 Mar 12	HG

03/12/12 A-P

Line	Conc.	Units	SD/RSD		1	2	3	4	5

***	Sample ID: ICV				Seq:	7	10:06:51	12 Mar 12	HG
	ICV								
Hg	3.71	ppb	.000		3.71				
***	Sample ID: ICB				Seq:	8	10:08:53	12 Mar 12	HG
	ICB								
Hg	-.129	ppb	.000		-.129				
***	Check Standard: 2 Ck2				Seq:	9	10:11:03	12 Mar 12	HG
Line	Flag	%Rcv.	Found	True	Units	SD/RSD			
Hg		105.	5.27	5.00	ppb	.000			
***	Check Standard: 1 Ck1				Seq:	10	10:13:03	12 Mar 12	HG
Line	Flag	Found	Range(+/-)	Units	SD/RSD				
Hg		-.116	.200	ppb	.000				
***	Sample ID: CRA				Seq:	11	10:15:31	12 Mar 12	HG
	CRA								
Hg	.122	ppb	.000		.122				
***	Sample ID: HIGH STD				Seq:	12	10:17:48	12 Mar 12	HG
	HIGH STD								
Hg	10.9	ppb	.000		10.9				
***	Sample ID: CHK STD				Seq:	13	10:19:51	12 Mar 12	HG
	CHK STD								
Hg	6.99	ppb	.000		6.99				
***	Sample ID: PB61618BL				Seq:	14	10:22:22	12 Mar 12	HG
	PBW								
Hg	-.098	ppb	.000		-.098				
***	Sample ID: PB61618BS				Seq:	15	10:24:23	12 Mar 12	HG
	LCSW								
Hg	4.15	ppb	.000		4.15				
***	Sample ID: D1753-01				Seq:	16	10:26:33	12 Mar 12	HG
Hg	-.057	ppb	.000		-.057				
***	Sample ID: D1753-03				Seq:	17	10:28:55	12 Mar 12	HG
Hg	-.080	ppb	.000		-.080				
***	Sample ID: D1753-03D				Seq:	18	10:31:16	12 Mar 12	HG
Hg	-.070	ppb	.000		-.070				

Line	Conc.	Units	SD/RSD	1	2	3	4	5

***	Check Standard: 2	Ck2		Seq:	19	10:33:48	12 Mar	12 HG
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		107.	5.34	5.00	ppb	.000		
***	Check Standard: 1	Ck1		Seq:	20	10:35:57	12 Mar	12 HG
Line	Flag	Found	Range(+/-)	Units	SD/RSD			
Hg		-.095	.200	ppb	.000			
***	Sample ID: D1753-03L			Seq:	21	10:38:21	12 Mar	12 HG
Hg	-.089	ppb	.000	-.089				
***	Sample ID: D1753-03S			Seq:	22	10:40:21	12 Mar	12 HG
Hg	1.35	ppb	.000	1.35				
***	Sample ID: D1753-03SD			Seq:	23	10:42:43	12 Mar	12 HG
Hg	1.34	ppb	.000	1.34				
***	Sample ID: D1753-03A			Seq:	24	10:44:47	12 Mar	12 HG
Hg	1.36	ppb	.000	1.36				
***	Sample ID: D1754-01			Seq:	25	10:47:06	12 Mar	12 HG
Hg	-.130	ppb	.000	-.130				
***	Sample ID: D1755-01			Seq:	26	10:49:09	12 Mar	12 HG
Hg	.075	ppb	.000	.075				
***	Check Standard: 2	Ck2		Seq:	27	10:51:13	12 Mar	12 HG
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		108.	5.39	5.00	ppb	.000		
***	Check Standard: 1	Ck1		Seq:	28	10:53:12	12 Mar	12 HG
Line	Flag	Found	Range(+/-)	Units	SD/RSD			
Hg		-.103	.200	ppb	.000			