

MERCURY ANALYSIS LOGBOOK

[illegible]

QA Control # A3040930

Page 54


MERCURY ANALYSIS LOGBOOK

[illegible]

QA Control # A3040930

Page 55

LB59716 A-P.



Instrument ID: A2 LB59716 A.P.

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1			<i>50</i>	Seq: 1		12:35:58 14 Mar 12	HG	
Hg	.000	ppb	9680					
*** Standard: 2 Rep: 1			<i>50.2</i>	Seq: 2		12:38:11 14 Mar 12	HG	
Hg	.200	ppb	27222					
*** Standard: 3 Rep: 1			<i>52.5</i>	Seq: 3		12:40:35 14 Mar 12	HG	
Hg	2.50	ppb	171960					
*** Standard: 4 Rep: 1			<i>55.0</i>	Seq: 4		12:42:36 14 Mar 12	HG	
Hg	5.00	ppb	323469					
*** Standard: 5 Rep: 1			<i>57.5</i>	Seq: 5		12:44:46 14 Mar 12	HG	
Hg	7.50	ppb	482714					
*** Standard: 6 Rep: 1			<i>510.0</i>	Seq: 6		12:47:26 14 Mar 12	HG	
Hg	10.0	ppb	624286					

03/14/12 A.P.

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: ICV								
			ICV	Seq: 7		12:50:04 14 Mar 12		HG
Hg	4.19	ppb	.000	4.19				
*** Sample ID: ICB								
			ICB	Seq: 8		12:52:06 14 Mar 12		HG
Hg	-.080	ppb	.000	-.080				
*** Check Standard: 2 Ck2								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		101.	5.05	5.00	ppb	.000		
*** Check Standard: 1 Ck1								
Line	Flag	Found	Range(+/-)	Units	Seq: 10	12:56:58 14 Mar 12		HG
Hg		-.021	.200	ppb	SD/RSD			
					.000			
*** Sample ID: CRA								
			CRA	Seq: 11		12:58:58 14 Mar 12		HG
Hg	.123	ppb	.000	.123				
*** Sample ID: HIGH STD								
			HIGH STD	Seq: 12		13:01:17 14 Mar 12		HG
Hg	10.8	ppb	.000	10.8				
*** Sample ID: CHK STD								
			CHK STD	Seq: 13		13:03:26 14 Mar 12		HG
Hg	6.01	ppb	.000	6.01				
*** Sample ID: PB61677BL								
			PBW	Seq: 14		13:05:40 14 Mar 12		HG
Hg	-.180	ppb	.000	-.180				
*** Sample ID: PB61677BS								
			LCSW	Seq: 15		13:07:53 14 Mar 12		HG
Hg	3.65	ppb	.000	3.65				
*** Sample ID: D1768-08								
				Seq: 16		13:10:16 14 Mar 12		HG
Hg	-.093	ppb	.000	-.093				
*** Sample ID: D1768-09								
				Seq: 17		13:12:21 14 Mar 12		HG
Hg	-.063	ppb	.000	-.063				
*** Sample ID: D1768-09D								
				Seq: 18		13:16:01 14 Mar 12		HG
Hg	-.052	ppb	.000	-.052				

Line	Conc.	Units	SD/RSD	1	2	3	4	5

***	Check Standard: 2 Ck2			Seq:	19	13:18:05 14 Mar 12 HG		
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		99.6	4.98	5.00	ppb	.000		
***	Check Standard: 1 Ck1			Seq:	20	13:20:39 14 Mar 12 HG		
Line	Flag	Found	Range(+/-)	Units	SD/RSD			
Hg		-.071	.200	ppb	.000			
***	Sample ID: D1768-09L			Seq:	21	13:23:18 14 Mar 12 HG		
Hg	-.079	ppb	.000	-.079				
***	Sample ID: D1768-10			Seq:	22	13:25:28 14 Mar 12 HG		
Hg	3.79	ppb	.000	3.79				
***	Sample ID: D1768-11			Seq:	23	13:27:28 14 Mar 12 HG		
Hg	3.58	ppb	.000	3.58				
***	Sample ID: D1768-09A			Seq:	24	13:30:36 14 Mar 12 HG		
Hg	4.25	ppb	.000	4.25				
***	Sample ID: D1768-12			Seq:	25	13:32:45 14 Mar 12 HG		
Hg	-.029	ppb	.000	-.029				
***	Sample ID: D1768-13			Seq:	26	13:34:45 14 Mar 12 HG		
Hg	-.028	ppb	.000	-.028				
***	Sample ID: D1768-14			Seq:	27	13:37:15 14 Mar 12 HG		
Hg	-.129	ppb	.000	-.129				
***	Sample ID: D1768-15			Seq:	28	13:40:07 14 Mar 12 HG		
Hg	-.188	ppb	.000	-.188				
***	Check Standard: 2 Ck2			Seq:	29	13:42:11 14 Mar 12 HG		
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		96.5	4.82	5.00	ppb	.000		
***	Check Standard: 1 Ck1			Seq:	30	13:44:22 14 Mar 12 HG		
Line	Flag	Found	Range(+/-)	Units	SD/RSD			
Hg		-.152	.200	ppb	.000			

Line	Conc.	Units	SD/RSD	1	2	3	4	5

***	Sample ID: D1768-16			Seq:	31	13:46:23	14 Mar 12	HG
Hg	-.086	ppb	.000	-.086				
***	Sample ID: D1768-17			Seq:	32	13:48:35	14 Mar 12	HG
Hg	-.061	ppb	.000	-.061				
***	Sample ID: D1768-18			Seq:	33	13:50:40	14 Mar 12	HG
Hg	-.067	ppb	.000	-.067				
***	Sample ID: D1768-19			Seq:	34	13:52:54	14 Mar 12	HG
Hg	-.068	ppb	.000	-.068				
***	Sample ID: PB61681BL			Seq:	35	13:55:04	14 Mar 12	HG
Hg	-.075	ppb	.000	PBW	-.075			
***	Sample ID: PB61681BS			Seq:	36	13:57:14	14 Mar 12	HG
Hg	3.42	ppb	.000	LCSW	3.42			
***	Sample ID: PB61672EB			Seq:	37	14:00:18	14 Mar 12	HG
Hg	-.087	ppb	.000	FLUID#1(03/12/12) -.087				
***	Sample ID: D1779-03			Seq:	38	14:02:37	14 Mar 12	HG
Hg	-.097	ppb	.000	-.097				
***	Check Standard: 2 Ck2			Seq:	39	14:15:10	14 Mar 12	HG
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		95.1	4.76	5.00	ppb	.000		
***	Check Standard: 1 Ck1			Seq:	40	14:21:08	14 Mar 12	HG
Line	Flag	Found	Range(+/-)	Units	SD/RSD			
Hg		-.189	.200	ppb	.000			
***	Sample ID: D1779-03D			Seq:	41	14:24:48	14 Mar 12	HG
Hg	-.090	ppb	.000	-.090				
***	Sample ID: D1779-03L			Seq:	42	14:26:47	14 Mar 12	HG
Hg	-.072	ppb	.000	-.072				

Line	Conc.	Units	SD/RSD	1	2	3	4	5	

***	Sample ID: D1779-03S			Seq:	43	14:28:49 14 Mar 12			HG
Hg	3.22	ppb	.000	3.22					
***	Sample ID: D1779-03SD			Seq:	44	14:30:51 14 Mar 12			HG
Hg	3.49	ppb	.000	3.49					
***	Sample ID: D1779-03A			Seq:	45	14:33:11 14 Mar 12			HG
Hg	4.26	ppb	.000	4.26					
***	Sample ID: D1779-04			Seq:	46	14:35:11 14 Mar 12			HG
Hg	-.113	ppb	.000	-.113					
***	Check Standard: 2 Ck2			Seq:	47	14:37:22 14 Mar 12			HG
Line	Flag	%Rcv.	Found	True	Units	SD/RSD			
Hg		98.1	4.91	5.00	ppb	.000			
***	Check Standard: 1 Ck1			Seq:	48	14:41:13 14 Mar 12			HG
Line	Flag	Found	Range(+/-)	Units	SD/RSD				
Hg		-.079	.200	ppb	.000				