

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD010220CLP.M
 Title : GC Extractables
 Last Update : Fri Jan 03 03:56:02 2020
 Response Via : Initial Calibration

Calibration Files

3 =PD056844.D 1 =PD056842.D 2 =PD056843.D
 4 =PD056845.D 5 =PD056846.D

	Compound	3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene	8.316	8.746	8.605	8.162	8.030	8.372	E5	3.57
2) A	alpha-BHC						0.000		-1.00
3) MA	gamma-BHC (Lindane)						0.000		-1.00
4) MA	Heptachlor						0.000		-1.00
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I						0.000		-1.00
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin						0.000		-1.00
14) MA	Endrin						0.000		-1.00
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD						0.000		-1.00
17) MA	4,4'-DDT						0.000		-1.00
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor						0.000		-1.00
21) B	Endrin ketone						0.000		-1.00
22)	Toxaphene-1	5.747	6.336	6.108	5.633	5.802	5.925	E3	4.88
23)	Toxaphene-2	6.231	6.615	6.448	6.188	6.306	6.358	E3	2.74
24)	Toxaphene-3	5.603	6.151	5.772	5.643	5.593	5.752	E3	4.07
25)	Toxaphene-4	1.559	1.625	1.587	1.606	1.644	1.604	E4	2.07
26)	Toxaphene-5	1.906	1.924	1.934	2.008	2.040	1.962	E4	2.97
27) SA	Decachlorobiphenyl	8.488	9.251	8.963	8.109	7.862	8.535	E5	6.76

Signal #2 Calibration Files

3 =PD056844.D 1 =PD056842.D 2 =PD056843.D
 4 =PD056845.D 5 =PD056846.D

	Compound	3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.510	1.577	1.557	1.489	1.464	1.519	E6	3.09
2) A	alpha-BHC						0.000		-1.00
3) MA	gamma-BHC (Lindane)						0.000		-1.00
4) MA	Heptachlor						0.000		-1.00
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I						0.000		-1.00
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin						0.000		-1.00
14) MA	Endrin						0.000		-1.00
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD						0.000		-1.00
17) MA	4,4'-DDT						0.000		-1.00
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor						0.000		-1.00
21) B	Endrin ketone						0.000		-1.00
22)	Toxaphene-1	8.763	8.887	8.132	8.414	8.495	8.538	E3	3.49
23)	Toxaphene-2	1.054	1.068	1.068	1.083	1.102	1.075	E4	1.70
24)	Toxaphene-3	1.530	1.585	1.534	1.537	1.522	1.542	E4	1.62

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD010220CLP.M
 Title : GC Extractables
 Last Update : Fri Jan 03 03:56:02 2020
 Response Via : Initial Calibration

Calibration Files

3 =PD056844.D 1 =PD056842.D 2 =PD056843.D
 4 =PD056845.D 5 =PD056846.D

	Compound	3	1	2	4	5	Avg		%RSD
25)	Toxaphene-4	4.268	4.189	4.219	4.325	4.240	4.248	E4	1.22
26)	Toxaphene-5	1.986	2.090	2.102	2.076	2.016	2.054	E4	2.45
27) SA	Decachlorobiphenyl	1.269	1.393	1.344	1.200	1.141	1.270	E6	8.09

(#) = Out of Range