

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD010324\
 Data File : PD080659.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jan 2024 12:18
 Operator : AR\AJ
 Sample : PEM082
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM094

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 01/04/2024
 Supervised By : Ankita Jodhani 01/04/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 03 21:33:23 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121623CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Dec 21 13:52:23 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.571	2.796	26634814	45367428	20.861	20.403
27)	SA Decachlor...	9.119	7.939	39513812	51175741	24.577	21.321
Target Compounds							
2)	A alpha-BHC	4.028	3.298	20625381	35323180	10.005	10.360
3)	MA gamma-BHC...	4.361	3.628	20460217	33558441	10.304	10.425
6)	B beta-BHC	4.555	3.924	9952088	16263586	9.588	9.472
14)	MA Endrin	6.616	5.662	76451706	131.6E6	48.810	50.010
16)	A 4,4'-DDD	6.748	5.810	213850	662972	0.154m	0.288m#
17)	MA 4,4'-DDT	7.063	6.061	156.1E6	250.2E6	105.600	103.756
18)	B Endrin al...	6.964	6.136	1442388	3635004	0.967	1.576 #
20)	A Methoxychlor	7.537	6.635	213.6E6	330.3E6	269.726	262.427
21)	B Endrin ke...	7.685	6.861	3895649	6673672	2.002m	2.086m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD010324\
Data File : PD080659.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jan 2024 12:18
Operator : AR\AJ
Sample : PEM082
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 03 21:33:23 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121623CLP.M
Quant Title : GC Extractables
QLast Update : Thu Dec 21 13:52:23 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

