

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD010721\
 Data File : PD060938.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Jan 2021 13:50
 Operator : DD\AJ
 Sample : PB133915BL
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 08 03:37:28 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD010721CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Jan 08 03:28:22 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.499	4.021	26651310	144.7E6	20.078	20.843
27)	SA Decachlor...	8.282	9.105	66366966	254.7E6	41.245	41.291
Target Compounds							
2)	A alpha-BHC	0.000	4.436	0	4113378	N.D.	0.386
3)	MA gamma-BHC...	0.000	4.663f	0	3091993	N.D.	0.316
6)	B beta-BHC	4.439f	4.912f	836469	2584379	1.028	0.651 #
7)	B delta-BHC	0.000	5.060	0	5925266	N.D.	0.627
8)	B Heptachlo...	5.155f	0.000	225107	0	0.125	N.D. #
15)	B Endosulfa...	6.286f	6.906f	4027084	1821713	2.817	0.293 #
16)	A 4,4'-DDD	0.000	6.906f	0	1821713	N.D.	0.312
24)	Toxaphene-3	6.286	0.000	4027084	0	218.397	N.D. #
26)	Toxaphene-5	7.096	0.000	4809510	0	107.444	N.D. #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD010721\
Data File : PD060938.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jan 2021 13:50
Operator : DD\AJ
Sample : PB133915BL
Misc :
ALS Vial : 23 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 08 03:37:28 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD010721CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jan 08 03:28:22 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm




Instrument :
ECD_D
ClientSampled :



#1 Tetrachloro-m-xylene


R.T.: 3.499 min
Delta R.T.: 0.002 min
Response: 26651310
Conc: 20.08 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.021 min
Delta R.T.: 0.000 min
Response: 144738506
Conc: 20.84 ng/ml




#2 alpha-BHC

R.T.: 0.000 min
Exp R.T. : 3.937 min
Response: 0
Conc: N.D.



#2 alpha-BHC


R.T.: 4.436 min
Delta R.T.: 0.024 min
Response: 4113378
Conc: 0.39 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 0.000 min
Exp R.T.: 4.223 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#3 gamma-BHC (Lindane)

R.T.: 4.663 min
Delta R.T.: -0.034 min
Response: 3091993
Conc: 0.32 ng/ml




#6 beta-BHC

R.T.: 4.439 min
Delta R.T.: -0.034 min
Response: 836469
Conc: 1.03 ng/ml



#6 beta-BHC

R.T.: 4.912 min
Delta R.T.: 0.047 min
Response: 2584379
Conc: 0.65 ng/ml



#7 delta-BHC

R.T.: 0.000 min
Exp R.T.: 4.680 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :

#7 delta-BHC


R.T.: 5.060 min
Delta R.T.: -0.024 min
Response: 5925266
Conc: 0.63 ng/ml

#8 Heptachlor epoxide

R.T.: 5.155 min
Delta R.T.: -0.070 min
Response: 225107
Conc: 0.12 ng/ml

#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T.: 5.902 min
Response: 0
Conc: N.D.



#15 Endosulfan II


R.T.: 6.286 min
Delta R.T.: -0.056 min
Response: 4027084
Conc: 2.82 ng/ml

Instrument :
ECD_D
ClientSampleId :




#15 Endosulfan II

R.T.: 6.906 min
Delta R.T.: -0.037 min
Response: 1821713
Conc: 0.29 ng/ml




#16 4,4'-DDD

R.T.: 0.000 min
Exp R.T. : 6.193 min
Response: 0
Conc: N.D.



#16 4,4'-DDD


R.T.: 6.906 min
Delta R.T.: 0.051 min
Response: 1821713
Conc: 0.31 ng/ml



#24 Toxaphene-3


R.T.: 6.286 min
Delta R.T.: 0.030 min
Response: 4027084
Conc: 218.40 ng/ml

Instrument :
ECD_D
ClientSampleId :




#24 Toxaphene-3

R.T.: 0.000 min
Exp R.T. : 7.114 min
Response: 0
Conc: N.D.




#26 Toxaphene-5

R.T.: 7.096 min
Delta R.T.: -0.016 min
Response: 4809510
Conc: 107.44 ng/ml



#26 Toxaphene-5


R.T.: 0.000 min
Exp R.T. : 7.691 min
Response: 0
Conc: N.D.



#27 Decachlorobiphenyl

R.T.: 8.282 min
Delta R.T.: 0.001 min
Response: 66366966
Conc: 41.24 ng/ml

Instrument :
ECD_D
ClientSampleId :



#27 Decachlorobiphenyl

R.T.: 9.105 min
Delta R.T.: 0.000 min
Response: 254653443
Conc: 41.29 ng/ml