

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD010722\
 Data File : PD067636.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Jan 2022 11:47
 Operator : AR\AJ
 Sample : N1031-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 BI-WEEKLY-DECANT

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 01/10/2022
 Supervised By :Ankita Jodhani 01/10/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 10 04:50:11 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PD122821.M
 Quant Title : GC Extractables
 QLast Update : Tue Dec 28 03:00:50 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.433	3.959	30723334	75985979	16.907m	11.516m#
2) SA Decachlor...	8.153	9.014	9012403	45701959	10.146m	9.616
Target Compounds						
5) MB Aldrin	4.687	5.435	21092975	51637135	7.621	6.964m
14) MA Endrin	5.968	6.673	18319765	61905644	8.721m	11.167m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD010722\
Data File : PD067636.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jan 2022 11:47
Operator : AR\AJ
Sample : N1031-01
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
BI-WEEKLY-DECANT


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 01/10/2022
Supervised By : Ankita Jodhani 01/10/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 10 04:50:11 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PD122821.M
Quant Title : GC Extractables
QLast Update : Tue Dec 28 03:00:50 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.433 min
Delta R.T.: -0.001 min
Response: 30723334
Conc: 16.91 ng/ml

Instrument :
ECD_D
ClientSampleId :
BI-WEEKLY-DECANT


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/10/2022
Supervised By :Ankita Jodhani 01/10/2022




#1 Tetrachloro-m-xylene

R.T.: 3.959 min
Delta R.T.: -0.005 min
Response: 75985979
Conc: 11.52 ng/ml m




#5 Aldrin

R.T.: 4.687 min
Delta R.T.: -0.003 min
Response: 21092975
Conc: 7.62 ng/ml



#5 Aldrin

R.T.: 5.435 min
Delta R.T.: -0.010 min
Response: 51637135
Conc: 6.96 ng/ml m




#14 Endrin

R.T.: 5.968 min
Delta R.T.: 0.003 min
Response: 18319765
Conc: 8.72 ng/ml

Instrument :
ECD_D
ClientSampleId :
BI-WEEKLY-DECANT


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/10/2022
Supervised By :Ankita Jodhani 01/10/2022




#14 Endrin

R.T.: 6.673 min
Delta R.T.: 0.004 min
Response: 61905644
Conc: 11.17 ng/ml m



#28 Decachlorobiphenyl

R.T.: 8.153 min
Delta R.T.: -0.004 min
Response: 9012403
Conc: 10.15 ng/ml m



#28 Decachlorobiphenyl

R.T.: 9.014 min
Delta R.T.: -0.005 min
Response: 45701959
Conc: 9.62 ng/ml