

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD010820\  
 Data File : PD056876.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 08 Jan 2020 11:13  
 Operator : AJ\MA  
 Sample : PIBLK04  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 ECD\_D  
 ClientSampleId :  
 PIBLK04

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Jan 09 00:41:37 2020  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD010220CLP.M  
 Quant Title : GC Extractables  
 QLast Update : Fri Jan 03 06:06:42 2020  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 µl  
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

|                             | Compound        | RT#1  | RT#2  | Resp#1   | Resp#2   | ng/ml  | ng/ml  |
|-----------------------------|-----------------|-------|-------|----------|----------|--------|--------|
| -----                       |                 |       |       |          |          |        |        |
| System Monitoring Compounds |                 |       |       |          |          |        |        |
| 1)                          | SA Tetrachlo... | 3.286 | 3.961 | 16543049 | 26693021 | 19.353 | 17.195 |
| 27)                         | SA Decachlor... | 7.959 | 8.998 | 32121097 | 43906764 | 32.959 | 30.229 |


Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD010820\  
Data File : PD056876.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 08 Jan 2020 11:13  
Operator : AJ\MA  
Sample : PIBLK04  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Jan 09 00:41:37 2020  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD010220CLP.M  
Quant Title : GC Extractables  
QLast Update : Fri Jan 03 06:06:42 2020  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Instrument :  
ECD\_D  
ClientSampled :  
PIBLK04