

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD010920\
 Data File : PD056887.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jan 2020 17:46
 Operator : AJ\MA
 Sample : INDA101
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA101

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 10 06:04:15 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD010920CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Jan 10 06:00:03 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.285	3.960	4157046	7503113	5.101	5.050
27)	SA Decachlor...	7.953	8.992	10164277	16278073	10.694	11.292
Target Compounds							
2)	A alpha-BHC	3.711	4.346	5593133	9565344	4.808	4.577
3)	MA gamma-BHC...	3.986	4.628	6303780	9597849	5.507	4.751
4)	MA Heptachlor	4.271	5.134	5357154	10431277	4.984	4.994
9)	A Endosulfan I	5.277	6.176	4543251	8907390	4.946	5.118
13)	MA Dieldrin	5.516	6.432	9925665	18181109	9.614	9.599
14)	MA Endrin	5.769	6.649	8546703	15062284	9.801	9.769
16)	A 4,4'-DDD	5.903	6.771	7343363	13617777	9.522	9.523
17)	MA 4,4'-DDT	6.138	7.069	5724405	12899500	7.821	9.050
20)	A Methoxychlor	6.687	7.526	19422107	40144618	48.648	51.373


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD010920\
Data File : PD056887.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jan 2020 17:46
Operator : AJ\MA
Sample : INDA101
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
INDA101

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 10 06:04:15 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD010920CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jan 10 06:00:03 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.285 min
Delta R.T.: 0.000 min
Response: 4157046
Conc: 5.10 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDA101




#1 Tetrachloro-m-xylene

R.T.: 3.960 min
Delta R.T.: 0.000 min
Response: 7503113
Conc: 5.05 ng/ml




#2 alpha-BHC

R.T.: 3.711 min
Delta R.T.: 0.000 min
Response: 5593133
Conc: 4.81 ng/ml



#2 alpha-BHC


R.T.: 4.346 min
Delta R.T.: 0.000 min
Response: 9565344
Conc: 4.58 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 3.986 min
Delta R.T.: 0.000 min
Response: 6303780
Conc: 5.51 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDA101




#3 gamma-BHC (Lindane)

R.T.: 4.628 min
Delta R.T.: 0.000 min
Response: 9597849
Conc: 4.75 ng/ml




#4 Heptachlor

R.T.: 4.271 min
Delta R.T.: 0.000 min
Response: 5357154
Conc: 4.98 ng/ml



#4 Heptachlor


R.T.: 5.134 min
Delta R.T.: 0.000 min
Response: 10431277
Conc: 4.99 ng/ml



#9 Endosulfan I


R.T.: 5.277 min
Delta R.T.: 0.000 min
Response: 4543251
Conc: 4.95 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDA101




#9 Endosulfan I

R.T.: 6.176 min
Delta R.T.: 0.000 min
Response: 8907390
Conc: 5.12 ng/ml




#13 Dieldrin

R.T.: 5.516 min
Delta R.T.: 0.000 min
Response: 9925665
Conc: 9.61 ng/ml



#13 Dieldrin


R.T.: 6.432 min
Delta R.T.: 0.001 min
Response: 18181109
Conc: 9.60 ng/ml



#14 Endrin


R.T.: 5.769 min
Delta R.T.: 0.000 min
Response: 8546703
Conc: 9.80 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDA101




#14 Endrin

R.T.: 6.649 min
Delta R.T.: 0.000 min
Response: 15062284
Conc: 9.77 ng/ml




#16 4,4'-DDD

R.T.: 5.903 min
Delta R.T.: 0.000 min
Response: 7343363
Conc: 9.52 ng/ml



#16 4,4'-DDD


R.T.: 6.771 min
Delta R.T.: 0.001 min
Response: 13617777
Conc: 9.52 ng/ml



#17 4,4'-DDT


R.T.: 6.138 min
Delta R.T.: 0.000 min
Response: 5724405
Conc: 7.82 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDA101




#17 4,4'-DDT

R.T.: 7.069 min
Delta R.T.: 0.000 min
Response: 12899500
Conc: 9.05 ng/ml




#20 Methoxychlor

R.T.: 6.687 min
Delta R.T.: 0.000 min
Response: 19422107
Conc: 48.65 ng/ml



#20 Methoxychlor


R.T.: 7.526 min
Delta R.T.: 0.000 min
Response: 40144618
Conc: 51.37 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.953 min
Delta R.T.: 0.000 min
Response: 10164277
Conc: 10.69 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDA101



#27 Decachlorobiphenyl

R.T.: 8.992 min
Delta R.T.: 0.000 min
Response: 16278073
Conc: 11.29 ng/ml