

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD010920\
 Data File : PD056887.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jan 2020 17:46
 Operator : AJ\MA
 Sample : INDA101
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA101

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 10 06:04:15 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD010920CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Jan 10 06:00:03 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.285	3.960	4157046	7503113	5.101	5.050
27)	SA Decachlor...	7.953	8.992	10164277	16278073	10.694	11.292
Target Compounds							
2)	A alpha-BHC	3.711	4.346	5593133	9565344	4.808	4.577
3)	MA gamma-BHC...	3.986	4.628	6303780	9597849	5.507	4.751
4)	MA Heptachlor	4.271	5.134	5357154	10431277	4.984	4.994
9)	A Endosulfan I	5.277	6.176	4543251	8907390	4.946	5.118
13)	MA Dieldrin	5.516	6.432	9925665	18181109	9.614	9.599
14)	MA Endrin	5.769	6.649	8546703	15062284	9.801	9.769
16)	A 4,4'-DDD	5.903	6.771	7343363	13617777	9.522	9.523
17)	MA 4,4'-DDT	6.138	7.069	5724405	12899500	7.821	9.050
20)	A Methoxychlor	6.687	7.526	19422107	40144618	48.648	51.373

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD010920\
Data File : PD056887.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jan 2020 17:46
Operator : AJ\MA
Sample : INDA101
Misc :
ALS Vial : 9 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 10 06:04:15 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD010920CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jan 10 06:00:03 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

